
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/9/17 8:19:07 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0509201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/9/17 8:02:59 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.9 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6164E-001+/- 0.012]	9.5453E-001	-6.0078E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3921E+000+/- 0.139]	2.2733E+000	-8.5634E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9182E+000+/- 0.086]	1.8783E+000	-4.6561E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6426E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2197E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4086E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4729E-001+/- 0.013]	5.4250E-001	-3.7810E-001 < : : : >
Corrected activity (964. [SD: 4.8428E-001+/- 0.007]	4.8442E-001	1.9376E-002 < : : : >
Corrected activity (1407 [SD: 4.8523E-001+/- 0.007]	4.8671E-001	2.0177E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)