



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
4/21/17 8:47:53 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 042120  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 12:00:00 AM  
Measurement Date: 4/21/17 7:49:33 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 351.8 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6184E-001+/- 0.013]	9.6654E-001	3.6201E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3730E+000+/- 0.130]	2.4682E+000	7.3386E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9115E+000+/- 0.079]	1.9122E+000	9.4646E-003 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6414E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2195E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4084E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4990E-001+/- 0.006]	5.4334E-001	-1.0204E+000 < : : : >
Corrected activity (964. [SD: 4.8420E-001+/- 0.008]	4.8470E-001	6.5529E-002 < : : : >
Corrected activity (1407 [SD: 4.8525E-001+/- 0.007]	4.8455E-001	-9.5247E-002 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)