



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
4/25/17 8:22:26 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: UNC Daily QC 042  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 12:00:00 AM  
Measurement Date: 4/25/17 7:58:28 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 351.4 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6196E-001+/- 0.013]	9.5164E-001	-8.0515E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3754E+000+/- 0.129]	2.1972E+000	-1.3822E+000 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9115E+000+/- 0.078]	1.7870E+000	-1.6059E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6434E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2198E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4087E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4974E-001+/- 0.006]	5.5027E-001	8.2825E-002 < : : : >
Corrected activity (964. [SD: 4.8422E-001+/- 0.007]	4.8911E-001	6.5675E-001 < : : : >
Corrected activity (1407 [SD: 4.8523E-001+/- 0.007]	4.8666E-001	1.9682E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)