
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/1/17 7:59:53 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0501201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/1/17 7:49:51 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 350.9 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6195E-001+/- 0.013]	9.5898E-001	-2.3769E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3787E+000+/- 0.132]	2.2657E+000	-8.5375E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9126E+000+/- 0.080]	1.8578E+000	-6.8860E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6435E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2199E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4088E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4950E-001+/- 0.006]	5.4671E-001	-4.4140E-001 < : : : >
Corrected activity (964. [SD: 4.8470E-001+/- 0.007]	4.8355E-001	-1.5564E-001 < : : : >
Corrected activity (1407 [SD: 4.8505E-001+/- 0.007]	4.8530E-001	3.5107E-002 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)