

LICENSEE EVENT REPORT

CONTROL BLOCK: (1)

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

0 1 N C B E P 2 0 0 - 0 0 0 0 0 - 0 0 3 4 1 1 1 1 4 5
7 8 9 14 15 25 26 30 37 CAT 58

CON'T
0 1 REPORT SOURCE L 0 5 0 - 0 3 2 4 7 0 8 3 1 7 9 8 0 9 2 5 7 9 9
7 8 60 61 68 69 74 75 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

0 2 During normal plant operation, the operator noticed that the drywell oxygen/hydrogen
0 3 monitor, CPC-1263, had a low flow condition and was not indicating normally. Drywell
0 4 oxygen concentration was maintained less than 4% by the CAC-1259 instrument.
0 5 Technical Specifications 3.6.6.4, 6.9.1.9b
0 6
0 7
0 8

0 9 SYSTEM CODE S F 11 CAUSE CODE E 12 CAUSE SUBCODE X 13 COMPONENT CODE A I R D R Y 14 COMP. SUBCODE Z 15 VALVE SUBCODE Z 16
7 8 9 10 11 12 13 18 19 20
17 LER/RO REPORT NUMBER 7 9 21 22 0 7 9 24 26 27 0 3 28 29 L 30 31 0 32
ACTION TAKEN X 18 FUTURE ACTION Z 19 EFFECT ON PLANT Z 20 SHUTDOWN METHOD Z 21 HOURS 0 0 0 0 22 ATTACHMENT SUBMITTED Y 23 NPRD-4 FORM SUB. Y 24 PRIME COMP. SUPPLIER A 25 COMPONENT MANUFACTURER H 0 4 1 26
33 34 35 36 37 40 41 42 43 44 47

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

1 0 An investigation found that the air dryer had frozen, blocking the sample air flow.
1 1 The dryer was deenergized and allowed to thaw, and a dry nitrogen purge was used to
1 2 blow out the moisture and dry the coils. The dryer was inspected by a contract heat-
1 3 ing and air-conditioning company and no problem could be found with the dryer. The
1 4 system was returned to service and monitored for several days with no problems noted;
7 8 9 (CON'T) 80

1 5 FACILITY STATUS E 28 % POWER 0 9 3 29 OTHER STATUS NA 30 METHOD OF DISCOVERY A 31 DISCOVERY DESCRIPTION Operator Surveillance 32
7 8 9 10 12 13 44 45 46 80
1 6 RELEASED OF RELEASE Z 33 Z 34 NA 35 AMOUNT OF ACTIVITY LOCATION OF RELEASE NA 36
7 8 9 10 11 44 45 80
1 7 PERSONNEL EXPOSURES NUMBER 0 0 0 37 TYPE Z 38 DESCRIPTION NA 39
7 8 9 11 12 13 80
1 8 PERSONNEL INJURIES NUMBER 0 0 0 40 DESCRIPTION NA 41
7 8 9 11 12 80
1 9 LOSS OF OR DAMAGE TO FACILITY TYPE Z 42 DESCRIPTION NA 43
7 8 9 10 80
PUBLCITY ISSUED N 44 DESCRIPTION NA 45
2 0 7 8 9 10 80

7909280427

1033 331 NRC USE ONLY

NAME OF PREPARER A. C. Tollison, Jr.

PHONE 919-457-9521

20 91-7-926

LER CONTINUATION -- RO# 2-79-079

Facility: BSEP Unit No. 2

Event Date: 8-31-79

therefore, this is considered an isolated event and no further action is required.