

# LICENSEE EVENT REPORT

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

CONTROL BLOCK: 1

1 G A E I H 2 7 0 0 0 0 0 0 0 0 0 0 3 4 1 1 1 1 4 5

LICENSEE CODE 14 15 LICENSE NUMBER 25 26 LICENSE TYPE 30 31 LAT 56 57

REPORT SOURCE L 0 5 0 0 0 3 6 6 7 0 9 1 5 7 9 8 1 0 0 3 7 9 9

DOCKET NUMBER 60 61 EVENT DATE 64 65 REPORT DATE 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES 10

2 The oxygen bottle providing the source of reagent gas for the hydrogen analyzers

3 (2P33-P001A and 2P33-P001B) was found to be empty. The empty bottle resulted in

4 the analyzers being inoperative. There was no effect on the environs. There have

5 been no similar occurrences of this type.

6

7

8

9

SYSTEM CODE S E 11 CAUSE CODE X 12 CAUSE SUBCODE X 13 COMPONENT CODE I N S T R U 14 COMP. SUBCODE X 15 VALVE SUBCODE Z 16

LER/RO REPORT NUMBER 17 EVENT YEAR 7 9 21 22 SEQUENTIAL REPORT NO. 1 0 4 24 26 OCCURRENCE CODE 0 3 28 29 REPORT TYPE L 30 REVISION NO. 0 32

ACTION TAKEN X 18 FUTURE ACTION X 19 EFFECT ON PLANT Z 20 SHUTDOWN METHOD Z 21 HOURS 0 0 0 0 22 ATTACHMENT SUBMITTED Y 23 NPRD-4 FORM SUB. N 24 PRIME COMP. SUPPLIER N 25 COMPONENT MANUFACTURER D 0 9 6 26

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS 27

10 The analyzers were inop because they had no reagent gas flow to the hydrogen cells. A

11 new bottle of gas was installed and the analyzers returned to operable status. A

12 standing order has been written to survey the usage of the reagent gases to determine

13 the frequency needed for replacing gas bottles. Procedure HNP-2-1060, Daily Rounds,

14 will be revised to include frequent checks of gas pressure. The Unit I system was

FACILITY STATUS E 28 % POWER 0 9 6 29 OTHER STATUS NA 30 METHOD OF DISCOVERY B 31 DISCOVERY DESCRIPTION Operator Observation 32

ACTIVITY CONTENT RELEASED OF RELEASE Z 33 Z 34 AMOUNT OF ACTIVITY NA 35 LOCATION OF RELEASE NA 36

PERSONNEL EXPOSURES NUMBER 0 0 0 37 TYPE Z 38 DESCRIPTION NA 39

PERSONNEL INJURIES NUMBER 0 0 0 40 DESCRIPTION NA 41

LOSS OF OR DAMAGE TO FACILITY TYPE Z 42 DESCRIPTION NA 43

PUBLICITY ISSUED N 44 DESCRIPTION NA 45

NAME OF PREPARER R. T. Nix PHONE: 912-367-7781

1138 255  
7910 10 387

NRC USE ONLY

Georgia Power Company  
Plant E. I. Hatch  
Baxley, Georgia

Reportable Occurrence Report No. 50-366/1979-104

Cause Description and Corrective Actions (cont.)

checked and found not to operate on bottled reagent gas.

1138 256