

**Strohmeyer, Daniel**

---

**From:** Strohmeyer, Daniel  
**Sent:** Thursday, May 23, 2019 6:21 AM  
**To:** kiayona@netnitco.net  
**Subject:** [ACTION] CN 611699 Indiana Regional PET Renewal Application

Kiayona Torres,

I am working on you renewal application for license 13-32724-01. There is substantial information missing from the renewal application and it cannot be process until all information is received. The guidance for medical license, NUREG-1556, Vol. 9, Rev. 2, provides all the information you need to be able to complete the application in full, including, but not limited to, a completed appendix C checklist, diagrams of your facility, and shielding calculations for applicable rooms. A link to the appropriate guidance is below.

<https://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>

Lastly, all application material must be signed by a member of management, i.e. CEO, etc. Please have the application form resigned when it is sent in.

Please provide all information by June 7<sup>th</sup>, so your application may be processed in a timely manner.

Please provide confirmation that this message has been received.

If you have any questions please contact me at your convenience.

Respectfully,

*Daniel C. Strohmeyer, CHP*

Health Physicist - Reviewer  
Materials Licensing Branch  
U.S. NRC, Region III  
630.829.9689 (O)  
630.515.1078 (F)