

Radiation Safety Office
106 Carrigan Drive, 004 Rowell Building
Burlington, Vermont 05405
(802) 656-2570
FAX: (802) 656-8876
radSAFE@uvm.edu

<https://www.uvm.edu/riskmanagement/radiation-safety>



The
UNIVERSITY
of VERMONT

August 16, 2019

Attn: Document Control Desk
U.S. Nuclear Regulatory Commission
Washington, DC 20555-0001

University of Vermont and State Agriculture College
License No. 44-00728-13
Docket No. 03013022

Reply to Notice of Violation

This letter is in response to the 'Notice of Violation' dated July 26, 2019 from the Nuclear Regulatory Commission. An NRC routine inspection took place at the University of Vermont on April 25, 2019 with one Severity Level IV violation identified. The following corrective actions will take place.

Routine radiation contamination surveys used to evaluate and report the amount of residual radioactivity will use the detector's efficiency. The contamination surveys will take into account the background counts, detector efficiency along with appropriate geometric factors in the determination of disintegrations per minute (dpm) or other appropriate radioactivity units.

Please contact me if you have any further questions regarding this matter.

Sincerely,


Thomas Kellogg, RSO

1007
RGN-1

cc: Regional Administrator, U.S. NRC, Region I