

DUKE POWER COMPANY

POWER BUILDING

422 SOUTH CHURCH STREET, CHARLOTTE, N. C. 28242

VONRO PPS

JUL 26 1979 04

WILLIAM O. PARKER, JR.
VICE PRESIDENT
STEAM PRODUCTION

July 23, 1979

TELEPHONE AREA 704
373-4083

Mr. James P. O'Reilly, Director
U. S. Nuclear Regulatory Commission
Region II
101 Marietta Street, Suite 3100
Atlanta, Georgia 30303

Re: RII:DMM
50-269/79-8
50-270/79-8
50-287/79-8

Dear Mr. O'Reilly:

Please refer to my letter of May 24, 1979 and the response to the infraction item of subject inspection report.

Mixed radionuclide sources prepared in the established physical volumes have been obtained for four of the eight counting configurations. Efficiencies for these configurations have been calculated and have been input to the operating GeLi detector systems.

It is expected that mixed radionuclide sources for outstanding configurations will be delivered and calibration completed by September, 1979.

Very truly yours,

William O. Parker, Jr.
William O. Parker, Jr. *By 7/26*

RLG:scs

908302

7909050019

~~721-267~~

790389

ORIGINAL COPY