

Gryglak, Magdalena

From: Gryglak, Magdalena
Sent: Wednesday, August 07, 2019 8:34 AM
To: pbyrne@mpcphysics.com
Cc: melissajennings@gmail.com
Subject: Request for additional information, NRC license no. 13-16518-01, Fayette Memorial Hospital, CN 613724
Attachments: nrc314.pdf

Good morning Mr. Byrne:

I have reviewed the licensee's request dated 7/10/19 to terminate the subject NRC license.

Please provide additional information:

- 1) Clarify the discrepancy between dates when the final status survey was conducted as provided in the letter dated 7/10/19 and the attached Close out Survey.
- 2) Provide the background contamination reading.
- 3) Provide the conversion factor (cpm to dpm) for the sodium iodide well counter, S/N 600155.
- 4) Provide your action levels for contamination and radiation readings.
- 5) Please explain the range of 0.0 to 0.02 mR/hr used when using the survey meter to perform the radiation survey (any readings above 0.02 mR/hr would not be measured)
- 6) Please confirm that you had one Co-57 flood source in your possession (two NRC Forms 540 were filled out for the same source).
- 7) Please provide signed copies of Form 540 (box 9) acknowledging that the waste/sources were received by the authorized recipient.
- 8) Provide completed NRC Form 314, "Certificate of Disposition of Materials," signed by duly authorized management representatives requesting termination of the above referenced license. The form can be found at: <https://www.nrc.gov/reading-rm/doc-collections/forms/nrc314.pdf> (form also attached).

Please provide your response in a signed and dated letter. You may email your response directly to me.

Please let me know if you have any questions.

Thank you

Magdalena R. Gryglak
US NRC Region III
630-829-9875