

From: Kennedy, Erin
To: Jasmine.forbes@arcelormittal.com
Subject: Additional Information Needed to Amend your Radioactive Materials License
Date: Wednesday, September 25, 2019 9:06:00 AM

Good Morning Jasmine,

Thank you for taking the time to discuss the amendment of your NRC license with me this morning. As we discussed, I will need some additional information to complete my review of your request.

In order to remove the Kay Ray gauges from your license I will need:

1. Disposal records showing that the gauges were transferred to an authorized entity
2. Final leak tests record for the gauges
3. A statement discussing the historical leak test data of the gauges (have you ever had a leaking gauge at your facility?)

Please provide this information in the form of a signed and dated letter.

In addition, you stated in your initial request that you only wanted to add one Berthold gauge at this time. If you include a request to add the gauge in your letter which includes the make model and specific activity of the second gauge, I can add that at this time as well. I do not need the serial number to add the gauge to the license.

Please let me know if you have any questions.

Thank you,

Erin Kennedy

U.S. Nuclear Regulatory Commission, Region III
Health Physicist- Materials Licensing
Email: erin.kennedy@nrc.gov
Office: 630-829-9876