

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Cardiac Diagnostic Center, LLC
10004 Kennerly Road
St. Louis, MO 63128

REPORT NUMBER(S) 2019-001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-38153

4. LICENSE NUMBER(S)

24-32766-01

5. DATE(S) OF INSPECTION

October 21, 2019

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
- (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Kevin G. Null		10/21/19
BRANCH CHIEF	Robert Ruiz		11/4/19

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Cardiac Diagnostic Center, LLC
10004 Kennerly Road
St. Louis, MO 63128

REPORT NUMBER(S) 2019-001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-38153

4. LICENSE NUMBER(S)

24-32766-01

5. DATE(S) OF INSPECTION

October 21, 2019

6. INSPECTION PROCEDURES USED

87130

7. INSPECTION FOCUS AREAS

All

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

2201

2. PRIORITY

5

3. LICENSEE CONTACT

A. Rashid Qureshi, M.D.

4. TELEPHONE NUMBER

(314) 609-6900



Main Office Inspection

Next Inspection Date: 10/21/2024



Field Office Inspection



Temporary Job Site Inspection

PROGRAM SCOPE

This was a routine, unannounced inspection of a nuclear medicine department that was authorized to use material described in 10 CFR 35.100 and 35.200. The licensee submitted a request for license termination on August 29, 2019. The request was being processed by the Region III licensing branch at the time of the inspection.

The inspector noted that the location of use was locked, and inaccessible. The inspector was able to locate property management and a staff member granted the inspector access to the licensee's location of use. The staff member escorted the inspector during a tour of the area.

The location of use was completely empty and vacated. The inspector examined all rooms, looked in cabinets and drawers, and conducted independent surveys. The inspector did not find any licensed material. Radiation levels were background, i.e., 0.02 - 0.05 mrem/hour.

No violations of NRC requirements were identified.