

Gryglak, Magdalena

From: Gryglak, Magdalena
Sent: Tuesday, October 01, 2019 12:35 PM
To: Laura T. Smith-; slaura@bronsonhg.org
Subject: Request to remove Cs-137 sources, NRC license no. 21-13125-01- Bronson Methodist Hospital

Ms. Speer-Smith,

I have reviewed your request dated August 22, 2019.

To continue my review, please provide the latest leak test results for the Cs-137 sealed sources in the self-contained irradiator before the sources were transferred.

Please provide a signed and dated letter transmitting the leak test results by 10/18/19.

You may provide the information directly to me via email.

Thank you

Magdalena R. Gryglak
US NRC Region III
630-829-9875