

Kelly, Jason

From: Kelly, Jason
Sent: Monday, December 9, 2019 7:36 AM
To: kfoli@hepn.com
Subject: U.S. NRC RML #13-18945-01

Mr. Foli:

Thank you for the information you provided in response to my request.

Your license also authorizes possession of Thermo Process Instruments, LP (formerly Texas Nuclear Corporation) Model 5190 Source Holders. Do you have an inventory of these devices at your facility?

Please clarify if you would like to maintain authorization to possess the Thermo Model 5190 Source Holders. As these devices continue to be manufactured/distributed by Thermo Process Instruments, LP, even if you do not currently have any of these in inventory currently you may wish to maintain authorization for these devices so as to be able to replace your Kay-Ray Model 7062P Source Holders that may be approaching the maximum recommended working life.

You may respond to me via facsimile. As a note, I am in training most of the week and will not be immediately available to respond if you should have any questions.

Jason M Kelly, MPH
Health Physicist
Materials Licensing Branch
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352
Office: (630) 829-9737
Fax: (630) 515-1078



<http://www.nrc.gov/>