



CONVERSATION RECORD

NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU		DATE OF CONTACT		TYPE OF CONVERSATION	
Jacqueline Katz, RT, MBA		11/21/2019		<input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING	
E-MAIL ADDRESS		TELEPHONE NUMBER			
jpkatz@comhs.org		219-836-7351			
ORGANIZATION		DOCKET NUMBER(S)			
The Community Hospital		030-09964			
LICENSE NAME AND NUMBER(S)		MAIL CONTROL NUMBER(S)			
The Community Hospital 13-15882-01		616051			
SUBJECT					
Pending NRC License Amendment Request - Additional Information Required					
SUMMARY AND ACTION REQUIRED (IF ANY)					
<p>This is a record of the conversation between Laura Cender and Jacqueline Katz of The Community Hospital that occurred on Nov. 21, 2019 regarding the license amendment request dated Oct. 1, 2019.</p> <p>This is in reference to the request to add Dr. Joseph Danavi, M.D. as a 10 CFR 35.200 authorized user. Specifically the letter supplied (dated Sept. 28, 2018) that was signed by Dr. Muhyaldeen Dia, M.D. and that appears to have been written to the Certification Board of Nuclear Cardiology does not supply the information required by 10 CFR 35.290.</p> <p>If Dr. Danavi is not a diplomate of the Certification Board of Nuclear Cardiology or has never been authorized on an NRC or Agreement State license as a authorized user of 10 CFR 35.200 then please provide the following information per the requirements of 10 CFR 35.290(c)(1)</p> <p>a.) Demonstration that Dr. Danavi has completed at least 700 hours of training and experience, including 80 hours of classroom and laboratory training.</p> <p>Classroom training must include training in the following areas:</p> <ul style="list-style-type: none">- Radiation physics and instrumentation;- Radiation protection;- Mathematics pertaining to the use and measurement of radioactivity;					
NAME OF PERSON DOCUMENTING CONVERSATION					
Laura B. Cender					
SIGNATURE				DATE OF SIGNATURE	
<i>Laura B. Cender</i>				11/21/2019	

CONVERSATION RECORD (continued)

LICENSE NAME AND NUMBER(S)

The Community Hospital
13-15882-01

MAIL CONTROL NUMBER(S)

616051

SUMMARY AND ACTION REQUIRED (IF ANY) (Continued)

- Chemistry of byproduct material for medical use; and
- Radiation Biology;

Work experience, under the supervision of an authorized user, must involve:

- Ordering, receiving, and unpacking radioactive materials safely and performing the related radiation surveys;
- Performing quality control procedures on instruments used to determine the activity of dosages, and performing checks for proper operation of survey meters;
- Calculating, measuring, and safely preparing patient or human research subject dosages;
- Using administrative controls to prevent a medical event involving the use of unsealed byproduct material;
- Using procedures to contain spilled byproduct material safely and using proper decontamination procedures;
- Eluting generator systems appropriate for preparation of radioactive drugs for imaging and localization studies, measuring and testing the eluate for radionuclidic purity, and processing the eluate with reagent kits to prepare labeled radioactive drugs; and
- Administering dosages of radioactive drugs to patients

Written attestation must be provided for Dr. Danavi confirming that he has satisfactorily completed the training required in 10 CFR 35.290(c)(1) and is able to independently fulfill the radiation safety-related duties as an authorized user for medical uses authorized under 10 CFR 35.200. This attestation may be provided by either an appropriate authorized user or residency program director as permitted by 10 CFR 35.290.

Please submit your signed and dated response by no later than December 13, 2019.

Cender, Laura

From: Cender, Laura
Sent: Thursday, November 21, 2019 1:25 PM
To: jpkatz@comhs.org
Subject: Pending Amendment of NRC License No. 13-15882-01 - Additional Information Requested
Attachments: Conversation to The Community Hospital - Nov. 21, 2019.pdf

Hello Jackie,

Thank you for taking time out of your day to discuss your pending amendment request for NRC License No. 13-15882-01. As we discussed, a record of our conversation detailing the additional information required to list Dr. Joseph Danavi, M.D. as an authorized user is attached.

Please submit your signed and dated response by no later than December 13, 2019. You may submit your response to our regional office via fax to 630-515-1078 or to our regional office via regular mail.

Please feel free to contact me at 630-829-9712 or send me an email if you have any questions.

Thank you,
Laura Cender

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712