

LICENSEE EVENT REPORT

CONTROL BLOCK: 1

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 G A E I H 1 2 0 0 - 0 0 0 0 0 0 - 0 0 3 4 1 1 1 1 1 4 5  
7 8 9 14 15 25 26 30 57 CAT 58

CON'T  
01 REPORT SOURCE L 6 0 5 0 0 0 3 2 1 7 0 6 2 4 8 1 8 0 7 1 5 8 1 9  
7 8 60 61 68 69 74 75 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES 10

02 While the plant was in steady state operation at 608 MWe, 1D11-K619B,  
03 Reactor Building Stack Monitor "B" was found to be inoperative. Technical  
04 Specification Section 3.2.8.3 requires both channels to be operative.  
05 Redundant monitor 1D11-K619A was operable. Plant operations was not  
06 affected as a result of this event. The health and safety of the public  
07 was not affected. This is a non-repetitive event.

08 \_\_\_\_\_  
7 8 9 80

09 SYSTEM CODE M B 11 CAUSE CODE E 12 CAUSE SUBCODE X 13 COMPONENT CODE I N S T R U 14 COMP. SUBCODE I 15 VALVE SUBCODE Z 16  
7 8 9 10 11 12 13 18 19 20  
17 LER/RO REPORT NUMBER 8 1 21 EVENT YEAR 8 1 22 SEQUENTIAL REPORT NO. 0 6 0 26 OCCURRENCE CODE 0 3 29 REPORT TYPE 1 30 REVISION NO. 0 32  
7 8 21 22 23 24 26 27 28 29 30 31 32  
ACTION TAKEN A 18 FUTURE ACTION Z 19 EFFECT ON PLANT Z 20 SHUTDOWN METHOD Z 21 HOURS 0 0 0 0 22 ATTACHMENT SUBMITTED Y 23 NRPD-4 FORM SUB. N 24 PRIME COMP. SUPPLIER N 25 COMPONENT MANUFACTURER V 1 1 5 26  
33 34 35 36 37 40 41 42 43 44 47

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS 27

10 The cause of the event has been attributed to component failure. The  
11 failure was due to broken photo multiplier tube. The component was  
12 replaced and monitor recalibrated and instrument returned to service.  
13 The unit is now in full compliance with the requirements.

14 \_\_\_\_\_  
7 8 9 80

15 FACILITY STATUS E 28 % POWER 0 9 6 29 OTHER STATUS NA 30 METHOD OF DISCOVERY A 31 DISCOVERY DESCRIPTION Lab Technician observation 32  
7 8 9 10 12 13 44 45 46 80

16 ACTIVITY RELEASED OF RELEASE Z 33 34 AMOUNT OF ACTIVITY NA 35 LOCATION OF RELEASE NA 36  
7 8 9 10 11 44 45 80

17 PERSONNEL EXPOSURES NUMBER 0 0 0 37 TYPE Z 38 DESCRIPTION NA 39  
7 8 9 11 12 13 80

18 PERSONNEL INJURIES NUMBER 0 0 0 40 DESCRIPTION NA 41  
7 8 9 11 12 80

19 LOSS OF OR DAMAGE TO FACILITY TYPE Z 42 DESCRIPTION NA 43  
7 8 9 10 80

20 PUBLICITY ISSUED N 44 DESCRIPTION 8107280407 810715 PDR ADDCK 05000321 S PDR 45 NRC USE ONLY  
7 8 9 10 58 59 80

NAME OF PREPARER R. T. Nix, Supt. of Maint.

PHONE 912-367-7781

LER No.: 50-321/1981-060  
Licensee: Georgia Power Company  
Facility: Edwin I. Hatch  
Docket No.: 50-321

Narrative Report  
for LER 50-321/1981-060

While the plant was in steady state operation at 608 MWe, 1D11-K619B, Reactor Building Stack Monitor "B", was found to be inoperative. Tech Specs Section 3.2.8.3 requires both channels to be operative. Redundant monitor 1D11-K619A was operable. Plant operation was not affected as a result of this event. The health and safety of the public was not affected. This is a non-repetitive event.

The cause of the event has been attributed to component failure. The failure was due to broken photo multiplier tube. The component was replaced and monitor recalibrated and instrument returned to service. The unit is now in full compliance with the requirements.

A generic review revealed no inherent problems.