

1	A	L	J	M	F	2	2	0	0	-	0	0	0	0	-	0	0	4	1	1	1	1	4		5				
5	9	LICENSEE CODE					14	15	LICENSE NUMBER										25	26	LICENSE TYPE					30	37	CAT 35	

REPORT SOURCE	L	6	0	5	0	0	0	3	5	4	7	0	6	0	3	8	1	3	0	7	0	7	8	1	3	
20	21	POCKET NUMBER										31	32	EVENT DATE					41	42	REPORT DATE					50

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

On 6-3-81, the "A" Recombiner Oxygen Outlet Analyzer was declared inoperable and removed from service when it was noted that the Recombiner Oxygen Analyzer was not responding to changes in oxygen concentration. Subsequent investigation revealed that the analyzer was in the "Calibrate" rather than the "Measure" mode of operation. Tech. Spec. 3.3.3.11, in part, requires the Oxygen Analyzer to be operable. Tech. Spec. 3.3.3.11 action statement requirements were met. Health/safety of the public was not affected.

SYSTEM CODE		CAUSE CODE		CAUSE SUBCODE		COMPONENT CODE						COMP SUBCODE		VALVE SUBCODE						
M	C	A		C		Z	Z	Z	Z	Z	Z	Z		Z						
9	10	11	12	13	14	15	16	17	18	19	20	21	22							
SERIAL REPORT NUMBER		EVENT YEAR		SEQUENTIAL REPORT NO.		OCCURRENCE CODE		REPORT TYPE		REVISION NO.										
17		8	1			0	2	2	/	0	3	L								
23	24	25	26	27	28	29	30	31	32	33	34									
ACTION TAKEN		FUTURE ACTION		EFFECT ON PLANT		SHUTDOWN METHOD		HOURS		ATTACHMENT SUBMITTED		NPD-4 FORM SLR		PRIME COMP. SUPPLIER		COMPONENT MANUFACTURER				
X	G	Z		Z		0	0	0	0	N		N		Z		Z	9	9	9	9
35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	52	53	54	55
CAUSE DESCRIPTION AND CORRECTIVE ACTIONS																				

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

The exact cause of this event has not been determined. However, the most probable cause was the failure of an I&C serviceman to place the Analyzer in the "Measure" mode following performance in Alignment Procedures. The Analyzer was placed in the "Measure" mode of operation and returned to service on 6-8-81 after several preventative maintenance tasks has been completed. A step will be added to FNP-2-IMP-223.15A requiring signoff to verify restoration of switches to proper position.

FACILITY STATUS	% POWER	OTHER STATUS	METHOD OF DISCOVERY	DISCOVERY DESCRIPTION
E	040	N/A	A	Operator Observation
ACTIVITY CONTENT RELEASED OR RELEASE	AMOUNT OF ACTIVITY	LOCATION OF RELEASE		
Z	N/A	N/A		
PERSONNEL EXPOSURES NUMBER TYPE DESCRIPTION				
000Z	N/A			
PERSONNEL INJURIES NUMBER DESCRIPTION				
000	N/A			
LOSS OF OR DAMAGE TO FACILITY TYPE DESCRIPTION				
Z	N/A			
PUBLICITY ISSUED DESCRIPTION				
N	N/A			

8107310363 810707

PDR ADOCK 05000364

S PDR

END PAGE 1