



CONVERSATION RECORD

NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU Ray A. Carlson, M.S.	DATE OF CONTACT 11/25/2019	TYPE OF CONVERSATION <input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING
E-MAIL ADDRESS rayacarlson@att.net	TELEPHONE NUMBER (734) 455-4730	
ORGANIZATION Medical Health Physics	DOCKET NUMBER(S) 030-39204	
LICENSE NAME AND NUMBER(S) Advanced Heart & Vascular Care 21-35571-01	MAIL CONTROL NUMBER(S) 616236	
SUBJECT Pending New NRC License Application - Additional Information Required		
SUMMARY AND ACTION REQUIRED (IF ANY) The is a summary of the conversation that occurred between Laura Cender and Ray A. Carlson on Nov. 21 2019 regarding the pending new NRC license application for Advanced Heart & Vascular Care dated Oct. 7, 2019. Please provide your appropriately signed and dated response to the following items by no later than Friday Dec. 20, 2019. You may submit your response via fax to 630-515-1078. 1. As Mr. Carlson is an outside consultant for Advanced Heart & Vascular Care who is requesting to be appointed on the license as the Radiation Safety Officer please provide the following information: - Identify other commitments of the consultant RSO for other NRC or Agreement State licensed facilities, along with a description of how Mr. Carlson will allocate time to permit the performance of the duties of the RSO as described in the regulations. State the minimum amount of time onsite (i.e.hours per week or days per quarter); - Identify an in-house representative who will serve as the point of contact during the RSO's absence. Include contact information such as phone and email; - Describe the overall availability of the consultant-RSO to respond to questions or operational issues that arise during the conduct of the radiation safety program and related regulatory requirements. - Specify the usual maximum amount of time that it will take for the consultant-RSO to arrive at the facility in the event of an emergency that requires their presence.		
NAME OF PERSON DOCUMENTING CONVERSATION Laura B. Cender		
SIGNATURE <i>Laura B. Cender</i>	DATE OF SIGNATURE 11/25/2019	

CONVERSATION RECORD (continued)

LICENSE NAME AND NUMBER(S)

Advanced Heart & Vascular Care
21-35571-01

MAIL CONTROL NUMBER(S)

616236

SUMMARY AND ACTION REQUIRED (IF ANY) (Continued)

2. Please provide the medical license number and issuing entity (e.g. state or territory) for Dr. Abdulwahhab Alroaini, M.D. and Dr. Ameen Abdulhamid Abdulmalik, M.D.
3. Please provide the following statement regarding training for individuals working in or frequenting restricted areas:

"We have developed and will implement and maintain written procedures for a program for training required under 10 CFR 19.12 for each group of workers, including (i) topics covered, (ii) qualifications of the instructors, (iii) method of training, (iv) method for assessing the success of the training, (v) initial training, and (vi) annual refresher training."
4. Please provide a facility diagram for your Lincoln Park, MI office that includes the following items:
 - a.) Specify which doors are access controlled (i.e. locked). Indicate if any other areas are access controlled.
 - b.) Please indicate the use of areas above and below locations where byproduct material is prepared, used, and stored.
5. Please provide one or both of the following statements regarding radiation monitoring instruments:
 - a.) "Radiation monitoring instruments will be calibrated by a vendor who is licensed by the NRC or an Agreement State to perform instrument calibrations."and/or
 - b.) "We have developed and will implement and maintain written radiation survey meter calibration procedures in accordance with the requirements in 10 CFR 20.1501 and that meet the requirements of 10 CFR 35.61."
6. Please provide one or both of the following statements regarding occupational dose:
 - a.) "We will maintain, for inspection by the NRC, documentation demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of the limits in 10 CFR 20.1502."or
 - b.) "We will monitor individuals in accordance with the criteria in the section titled, 'Radiation Safety Program—Occupational Dose' in NUREG-1556, Vol. 9, Rev. 3, 'Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Medical Use Licensees.'

CONVERSATION RECORD (continued)

LICENSE NAME AND NUMBER(S)

Advanced Heart & Vascular Care
21-35571-01

MAIL CONTROL NUMBER(S)

616236

SUMMARY AND ACTION REQUIRED (IF ANY) (Continued)

7. Please provide the following statement regarding material receipt and accountability:

- a.) "We will develop, implement, and maintain written procedures for licensed material accountability and control to ensure that:
- license possession limits are not exceeded
 - licensed material in storage is secured from unauthorized access or removal
 - licensed material not in storage is maintained under constant surveillance and control
 - records of receipt (either from the licensee's own production operations or from another licensee), transfer, and disposal of licensed material, are maintained."

8. Please provide the following statement regarding safe use of unsealed licensed material:

- a.) "We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR 20.1201."

9. Please provide the following statement regarding waste management:

- a.) "We will contact the appropriate NRC Regional Office for guidance on treatment or disposal of waste by incineration or compaction."

Cender, Laura

From: Cender, Laura
Sent: Monday, November 25, 2019 11:01 AM
To: rayacarlson@att.net
Subject: Pending New License Application for Advanced Heart & Vascular Care - Additional Information Required
Attachments: Conversation Record to Advanced Heart & Vascular Care. - Nov. 25, 2019.pdf

Hello Ray,

Thank you for taking time out of your morning to discuss the pending new license application for Advanced Heart & Vascular Care. As discussed, I have attached a record of our conversation detailing the additional information required.

Additionally, as we discussed this morning, a pre-licensing site visit will be required for this applicant. I will be available the week of January 6th, 2020 or alternatively I will also be available during the week between Christmas and New Year's. If possible, I would like to coordinate this site visit with the pre-licensing site visit for Zakwan Mahjoub, M.D., P.C. for convenience of travel arrangements due to the relatively close proximity of these two applicants.

Please let me know if you have any questions or concerns.

Respectfully,
Laura Cender

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712