

CONTROL BLOCK: | | | | | | | ① (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

CON'T

REPORT SOURCE	DOCKET NUMBER	EVENT DATE	REPORT DATE
01	6050003137	11226820	80112681

02 | At 1945 after returning containment (CTMT) air particulate (APD) and  
03 | gaseous monitors to service (Ref. LER 80-59 U-2) the APD remained in-  
04 | operable. The APD was repaired and returned to service at 1330 on  
05 | 12-30-80. Grab samples were taken prior to and during the event. CTMT  
06 | sum level alarm and gaseous monitor were operable during the event.  
07 | This has not been a repetitive event.

08		9		SYSTEM CODE		CAUSE CODE		CAUSE SUBCODE		COMPONENT CODE		COMP SUBCODE		VALVE SUBCODE	
7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22
0	8	3	R	E	F	I	N	S	T	R	U	X	Z		
7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22
LER/RO REPORT NUMBER		EVENT YEAR		SEQUENTIAL REPORT NO.		OCCURRENCE CODE		REPORT TYPE		REVISION NO.					
17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	32
2	0	2	0	0	6	0	0	3	L	0					
ACTION TAKEN		FUTURE ACTION		EFFECT ON PLANT		SHUTDOWN METHOD		HOURS		ATTACHMENT SUBMITTED		NPRD-4 FORM SUB.		PRIME COMP SUPPLIER	
33	34	35	36	37	38	39	40	41	42	43	44	45	46	47	48
A	Z	Z	Z	0	0	0	0	N	N	A					

1 0 The air particulate detector function box connector (Deutsch DS07-12S-059)

1 1 failed resulting in abnormal high indication. Connector wear from handl-

1 2 ing during preventive and corrective maintenance contributed to an end-

1 3 of-life type of failure with loosened connections at the rear of the con-

1 4 nector. No preventive action is deemed necessary at this time.

FACILITY STATUS		% POWER		OTHER STATUS		METHOD OF DISCOVERY		DISCOVERY DESCRIPTION	
1	5	E	28	0	6	0	29	NA	44
7	8	9	10	11	12	13	14	15	16
ACTIVITY CONTENT		RELEASED OF RELEASE		AMOUNT OF ACTIVITY		LOCATION OF RELEASE			
1	6	Z	33	Z	34	NA	44	45	46
7	8	9	10	11	12	13	14	15	16
PERSONNEL EXPOSURES		NUMBER		TYPE		DESCRIPTION			
1	7	0	0	0	37	Z	38	NA	44
7	8	9	10	11	12	13	14	15	16
PERSONNEL INJURIES		NUMBER		DESCRIPTION					
1	8	0	0	0	40	NA	44	45	46
7	8	9	10	11	12	13	14	15	16
LOSS OF OR DAMAGE TO FACILITY		TYPE		DESCRIPTION					
1	9	Z	42	NA	44	45	46	47	48
7	8	9	10	11	12	13	14	15	16
PUBLICITY		ISSUED		DESCRIPTION					
2	0	1	44	NA	44	45	46	47	48
7	8	9	10	11	12	13	14	15	16

NAME OF PREPARER S.M. Davis/P.G. Rizzo

PHONE: (301) 269-4742/4786