

PROCEDURE

PROCEDURE TITLE

HNP-8152

PROCEDURE NUMBER

Lab

RESPONSIBLE SECTION

NON-SAFETY RELATED ()

HNP-9

REFERENCE ONLY

MANUAL SET

WE
PROCEDURE REVIEW REQUEST
FOR NEW PROCEDURE

Recd 4/24

DRM
4/24/82

PROCEDURE NO. HNP- 8152

SAFETY RELATED ☒ 5/14/82 NON-SAFETY RELATED ☒

REQUESTED BY		DEPARTMENT HEAD APPROVAL	
Name:	Date:	Signature:	Date:
R. ANDERSON	3-30-82	W.H. Royn	3-31-82

Attach marked up copy of procedure to this form.

REASON FOR REQUEST A PROCEDURE IS NEEDED TO DEFINE
THE INTERVALS IN WHICH THE HEALTH PHYSICS
INSTRUMENTS ARE CALIBRATED.

PRB RECOMMENDS APPROVAL: ☒ YES ☐ NO

82-79
PRB Number

Steve Lipp
PRB Secretary

5/14/82
Date


REFERENCE ONLY

HNP-3

MANUAL SET

APPROVAL
See Title Page
DATE
See Title Page

E. I. HATCH NUCLEAR PLANT

Georgia Power 

PROCEDURE NO
HNP-8152
REVISION NO
0
PAGE NO
1 of 1

 CH

INSTRUMENT CALIBRATION INTERVALS

A. PURPOSE

Health Physics instruments are calibrated at various intervals. This procedure is to define the intervals in which the Health Physics instruments are calibrated and or checked.

B. REFERENCE

American National Standards ANSI N-323-1978 Radiation Protection Instrumentation Test and Calibration

C. DEFINITIONS

1. Annually - 365 days from the previous due date plus or minus 90 days.

FOR EXAMPLE:

Earliest date for cal. 1-1-82
Due date 4-1-82
Latest date for cal. 6-29-82

NOTE

All due dates are assigned to instruments via the Instrument Calibration Due Date Log Book.

2. Semiannually - 182 days from the previous due date plus or minus 45 days.
3. Quarterly - 92 days from the previous due date plus or minus 23 days.
4. Monthly - 30 days from the previous due date plus or minus 7 days.
5. Weekly - 7 days from the previous due date plus or minus 2 days.

REFERENCE ONLY

MANUAL SET

WE
 PROCEDURE REVIEW REQUEST
 FOR NEW PROCEDURE

RAM
 4/24/82

PROCEDURE NO. HNP- 8152

SAFETY RELATED ☒ *RAM* 5/14/82 NON-SAFETY RELATED ☒

REQUESTED BY		DEPARTMENT HEAD APPROVAL	
Name:	Date:	Signature:	Date:
R. ANDERSON	3-30-82	<i>W. H. Royce</i>	3-31-82

Attach marked up copy of procedure to this form.

REASON FOR REQUEST A PROCEDURE IS NEEDED TO DEFINE
THE INTERVALS IN WHICH THE HEALTH PHYSICS
INSTRUMENTS ARE CALIBRATED.

PRB RECOMMENDS APPROVAL: ☒ Yes ☐ No

Steve Lipp
 PRB Secretary

82-79
 PRB Number

5/14/82
 Date

HNP-3

REFERENCE ONLY

CH
 MANUAL SET