




KANSAS GAS AND ELECTRIC COMPANY

GLENN L. KOESTER  
VICE PRESIDENT - NUCLEAR

June 9, 1982

Mr. W.C. Seidle, Chief  
Reactor Projects Branch 2  
U.S. Nuclear Regulatory Commission,  
Region IV  
611 Ryan Plaza Drive, Suite 1000  
Arlington, Texas 76011



KMLNRC 82-211  
Re: Docket No. STN 50-482  
Ref: Letter KMLNRC 82-196 dated 5/4/82 from GLKoester,  
KG&E, to WCSeidle, NRC  
Subj: Interim 10CFR50.55(e) Report - Post Accident  
Instrument Errors

Dear Mr. Seidle:

This letter provides an interim report on a 10CFR50.55(e) matter concerning post accident errors in the reactor coolant wide range pressure instruments. This item is also the subject of Information Notice No. 82-11.

Westinghouse has determined that the post accident inaccuracies of the subject instruments constitute a significant deficiency because the design does not conform to the criteria and bases stated in our Final Safety Analysis Report. In particular, the accuracy functional requirements for Regulatory Guide 1.97, Revision 1, are not demonstrated to be satisfied by the current instrumentation.

A thorough evaluation of this issue is in progress to determine and analyze possible solutions. It is anticipated that it will be several months before this issue can be resolved for Wolf Creek Generating Station. The status of this evaluation will be reported to you on or before November 4, 1982.

In the interim, please direct any questions on this matter to me or to Mr. Otto Maynard of my staff.

Yours very truly,

*Glenn L. Koester*

15-27

B206150446 820609  
PDR ADOCK 05000482  
S PDR

GLK:bb  
cc: Mr. Thomas Vandel