

**From:** [Gallagher, Robert](#)  
**To:** [Regina Sheehan](#)  
**Subject:** Request for Additional Information - Mail Control No. 617690  
**Date:** Thursday, February 27, 2020 8:32:00 AM  
**Attachments:** [image002.png](#)

---

License No. 06-13905-01  
Docket No. 03001300  
Control No. 617690

PLEASE CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION BY RETURN EMAIL

Ms. Sheehan,

This is regarding your letter dated January 17, 2020 requesting an amendment to include new schematic diagrams of the areas of use within your facility. In order for us to continue our review of your request more information is necessary.

1. In the diagram labeled as "Original" the bathroom appears to be part of the restricted area within the Nuclear Medicine suite. Please submit radiological survey data verifying no radioactive contamination exists in the room (the bathroom).

We will continue our review upon receipt of the requested information. Please respond within 30 days of the date of this request. You may respond to my attention by letter, email (if the letter is signed by senior management and scanned into a pdf format), or fax (610-337-5269), referencing mail control number 617690. If you have any questions regarding this deficiency request, please contact me at (610) 337-5182.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Regards,

Robert L. Gallagher  
Health Physicist  
U.S. NRC, Region I  
2100 Renaissance Blvd.  
King of Prussia, PA 19406  
(610) 337-5182 office  
(610) 337-5269 fax

