

Hill, Carol

From: Michlig, Paul <pmichlig@amengtest.com>
Sent: Monday, March 09, 2020 10:41 AM
To: Hill, Carol
Subject: [External_Sender] RE: Acknowledgement of Receipt of Amendment Request is attached (005) (00B) (003).pdf
Attachments: Casper - Jan 2020 AET Leak Tests.pdf; Casper AET 2020 Refresher.pdf; Gillette - Jan 2020 AET Leak Tests.pdf; Gillette AET 2020 Refresher Certs.pdf

Hi Carol

We completed the leak tests for the gauges AET took ownership of from Strata and the Strata gauge operators also completed our AET annual nuclear gauge refresher, I attached this information for your review and your files.

Thanks,
Paul

Paul Michlig
Construction Services Manager
pmichlig@amengtest.com
D: 715.359.3534 | C: 715.571.7309



American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

www.amengtest.com

Please read our email disclaimer at this [link](#) on our website.

PUBLIC

- ☐ Immediate Release
☒ Normal Release

NON-PUBLIC

- ☐ A.3 Sensitive-Security Related
☐ A.7 Sensitive Internal
☐ Other: _____

Reviewer: MRS Date: 3/11/20

Notification

CCA

From: Hill, Carol <Carol.Hill@nrc.gov>
Sent: Monday, December 30, 2019 8:47 AM
To: Michlig, Paul <pmichlig@amengtest.com>
Subject: Acknowledgement of Receipt of Amendment Request is attached (005) (00B) (003).pdf

[Warning - External Source] Please be careful about responding to or opening any attachments or links in this message, as it did r

Carol L. Hill, Licensing Assistant

Direct: 817-200-1140
Toll Free: 1-800-952-9677
Fax: 817-200-1083
E-mail: Carol.Hill@nrc.gov

US Nuclear Regulatory Commission
1600 E. Lamar Blvd.
Arlington, TX 76011-4511

RECEIVED
MAR 09 2020

DNMS

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/23/2020
Test Number: 7

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE

NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3500
Gauge S/N: 779

Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
CZ1187	0.00022
066/07	0.00001

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.

Reviewed by:

Date:

01/23/2020

Customer Signature:

Date:

1/24/20

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

618280

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/22/2020
Test Number: 3

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE

NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge S/N: 25175

Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
47-21444	0.00023
75-7352	0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.^o The source(s) tested above may remain in use.

Reviewed by:

Date:

01/22/2020

Customer Signature:

Date:

1/24/20

^oCPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

618280

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/22/2020
Test Number: 7

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE

NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge S/N: 33500

Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
47-1597	0.00023
750-9053	0.00001

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.

Reviewed by:

Date:

01/22/2020

Customer Signature:

Date:

1/24/20

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

618280

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/22/2020
Test Number: 9

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE

NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge S/N: 33501

Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
47-2942	0.00019
750-9054	0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.* The source(s) tested above may remain in use.

Reviewed by:

Date:

01/22/2020

Customer Signature:

Date:

1/24/20

*CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/22/2020
Test Number: 6

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE
NC Materials License #092-1073-1


This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge S/N: 30148


Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
47-27300	0.00025
750-4426	0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.

Reviewed by: 

Date: 01/22/2020

Customer Signature: 

Date: 1/24/20

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

618280

InstroTek, Inc.
1 Triangle Drive, PO Box 13944
Research Triangle Park, NC 27709
(919)875-8371 Fax (919)875-8328

1/22/2020
Test Number: 6

STEVE JENSEN Phone: (307)234-2126
AMERICAN ENGINEERING TESTING, INC. (CASPER Fax: (307)266-5143
605 N. WAREHOUSE RD
CASPER, WY 82601

LEAK TEST CERTIFICATE
NC Materials License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.


Gauge Model: 3430
Gauge S/N: 226

Test Date: 1/17/2020

Source (Model/Serial#)	Reading in microCuries
47-27029	0.00019
750-4546	0.00000

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.° The source(s) tested above may remain in use.

Reviewed by:



Date:

01/22/2020

Customer Signature:



Date:

1/24/20

°CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137.

618280



CERTIFICATE OF COMPLETION

This certifies that

TREVOR OLSON

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date

618280



CERTIFICATE OF COMPLETION

This certifies that

COLTER CARZOLI

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date

618280



CERTIFICATE OF COMPLETION

This certifies that

BRYAN PETERSEN

HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date



CERTIFICATE OF COMPLETION

This certifies that

JARED JUNG

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020
Date

618280



CERTIFICATE OF COMPLETION

This certifies that

LOREN DRINKWALTER

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date

618280



CERTIFICATE OF COMPLETION

This certifies that

STEPHEN JENSEN

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date

618280



CERTIFICATE OF COMPLETION

This certifies that

DALLAS JAEHN

HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 8, 2020

Date

618280

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

JAMES FREY
AET
2801 E. 2ND ST
GILLETTE, WY 82718

Cust ID: 14289

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 25174

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-21443	02/28/1995	Am-241:Be	1.48	40
75-7351	05/03/1995	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 01/31/2020
Sample analyzed on: 02/06/2020 11:43:55 AM Position: 55
Analyzed by: RM

Reviewed by:

Date: 03-1-20

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.20E+01	1.92E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	28
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.3E-01	1.4E+00

This certifies that the leak test results are:

☒ Less than 185 Bq (0.005 uCi) ☐ Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

JAMES FREY
AET
2801 E. 2ND ST
GILLETTE, WY 82718

Cust ID: 14289

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 37327

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-4470	01/18/2006	Cs-137	0.296	8
78-2129	09/22/2005	Am-241:Be	1.48	40

Reviewed by: *[Signature]*
Date: 3-14-20

LEAK TEST ANALYSIS:

Sample collected on: 01/31/2020

Sample analyzed on: 02/06/2020 11:48:45 AM Position: 59

Analyzed by: RM

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.20E+01	1.92E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	22
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.3E-01	1.4E+00

This certifies that the leak test results are:

☒ Less than 185 Bq (0.005 uCi)

☐ Greater than 185 Bq (0.005 uCi)

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

JAMES FREY
AET
2801 E. 2ND ST
GILLETTE, WY 82718

Cust ID: 14289

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430

Serial No: 37328

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
78-2130	09/22/2005	Am-241:Be	1.48	40
77-4471	01/18/2006	Cs-137	0.296	8

Reviewed by:

LEAK TEST ANALYSIS:

Sample collected on: 01/31/2020

Sample analyzed on: 02/06/2020 11:46:20 AM

Position: 57

Analyzed by: RM

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.20E+01	1.92E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	14
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.3E-01	1.4E+00

This certifies that the leak test results are:☒ Less than 185 Bq (0.005 uCi)☐ Greater than 185 Bq (0.005 uCi)

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

JAMES FREY
AET
2801 E. 2ND ST
GILLETTE, WY 82718

Cust ID: 14289

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430

Serial No: 37334

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-4477	01/18/2006	Cs-137	0.296	8
78-2136	09/22/2005	Am-241:Be	1.48	40

Reviewed by:

Date: 3-6-20

LEAK TEST ANALYSIS:

Sample collected on: 01/31/2020

Sample analyzed on: 02/06/2020 11:47:32 AM

Position: 58

Analyzed by: RM

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.20E+01	1.92E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	27
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.3E-01	1.4E+00

This certifies that the leak test results are:☒ Less than 185 Bq (0.005 uCi)☐ Greater than 185 Bq (0.005 uCi)

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 878-9537 Fax: (866) 391-2759

License: NC 032-0182-1

JAMES FREY
AET
2801 E. 2ND ST
GILLETTE, WY 82718

Cust ID: 14289

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430

Serial No: 61955

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
78-5312	05/28/2008	Am-241:Be	1.48	40
77-8807	10/27/2008	Cs-137	0.296	8

Reviewed by: *[Signature]*
Date: 3-16-20

LEAK TEST ANALYSIS:

Sample collected on: 01/31/2020

Sample analyzed on: 02/06/2020 11:45:07 AM

Position: 56

Analyzed by: RM

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.20E+01	1.92E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	22
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.3E-01	1.4E+00

This certifies that the leak test results are:☒ Less than 185 Bq (0.005 uCi)☐ Greater than 185 Bq (0.005 uCi)



CERTIFICATE OF COMPLETION

This certifies that

JAMES FREY

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 22, 2020

Date

618280



CERTIFICATE OF COMPLETION

This certifies that

JASON GILLESPIE

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 20, 2020

Date

618282



CERTIFICATE OF COMPLETION

This certifies that

ROBERT COWAN

**HAS SUCCESSFULLY COMPLETED THE HAZMAT TRANSPORTATION
REFRESHER TRAINING AND ANNUAL REFRESHER COURSE
FOR THE USE OF NUCLEAR GAUGES**

Conducted by personnel from

American Engineering Testing, Inc.



Paul Michlig, RSO
American Engineering Testing, Inc.
4203 Schofield Avenue, Suite 1
Schofield, WI 54476

January 27, 2020

Date

618280



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee

Mr. Paul Michlig
Corporate Radiation Safety Officer
American Engineering Testing, Inc.
550 Cleveland Avenue
St. Paul, Minnesota 55114

Date

03/11/2020

License Number(s)

22-20271-01

Mail Control Number(s)

618280

Licensing and/or Technical Reviewer or Branch

E. Gilman

This is to acknowledge receipt of your: ☒ Letter and/or ☐ Application Dated: 01/23/2020

The initial processing, which included an administrative review, has been performed.

☒ Amendment ☐ Termination ☐ New License ☐ Renewal

☐ There were no administrative omissions identified during our initial review.

☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

☐ Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>
Follow the instructions on the form for submission.

☐ The following administrative omissions have been identified:

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region IV
U. S. Nuclear Regulatory Commission
DNMS/NMSB - B
1600 E. Lamar Boulevard
Arlington, TX 76011-4511
(817) 200-1103 or (817) 200-1140

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM WBL

Program Code: 03121
Status Code: Pending Amendment
Fee Category: 3P
Exp. Date: 09/30/2022
Fee Comments:
Decom Fin Assur Req: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: American Engineering Testing, Inc.
Received Date: 03/09/2020
Docket Number: 3018180
Mail Control Number: 618280
License Number: 22-20271-01
Action Type: Notifications

2. FEE ATTACHED

Amount: 

Check No.: 

3. COMMENTS

Signed: 

Date: 20200311

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____