


LICENSEE EVENT REPORT

CONTROL BLOCK: 

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

0	1	G	A	E	I	H	2	2	0	0	-	0	0	0	0	0	-	0	0	3	4	1	1	1	1	4			5		
7	8	LICENSEE CODE						14	15	LICENSE NUMBER										25	26	LICENSE TYPE					30	57	CAT		58

CON'T

REPORT SOURCE 0 1 7 8 L 6 0 5 0 0 0 3 6 6 7 0 2 1 4 8 3 8 0 3 1 0 8 3 9
60 61 DOCKET NUMBER 68 69 EVENT DATE 74 75 REPORT DATE 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

0 2 | During performance of the "COMSIP DELPHI MODEL K-1V H2 O2 ANALYZER"
0 3 | procedure, the worker discovered that Drywell H2 O2 "B" analyzer would
0 4 | not calibrate. It was declared inoperable. This event is contrary to
0 5 | Tech. Specs. section 3.3.6.4, and Table 3.3.6.4-1, Item 9. A 30-day LCO
0 6 | was established per Tech. Specs. section 3.3.6.4, ACTION a. Public
0 7 | health and safety were not affected by this repetitive event as last
0 8 | reported on LER 50-366/1982-063.

09		SYSTEM CODE S E 11		CAUSE CODE E 12		CAUSE SUBCODE B 13		COMPONENT CODE I N S T R U 14				COMP. SUBCODE E 15		VALVE SUBCODE Z 16	
7 8		9 10		11 12		12 13		13 18				19 20		20 21	
17 LER/RO REPORT NUMBER		EVENT YEAR 8 3 21 22		23		SEQUENTIAL REPORT NO. 0 0 3 24 26		27		OCCURRENCE CODE 0 3 28 29		REPORT TYPE L 30		31	
32		REVISION NO. 0 32		33		ACTION TAKEN A 18		FUTURE ACTION F 19		EFFECT ON PLANT Z 20 35		SHUTDOWN METHOD Z 21 36		HOURS 0 0 0 0 22 37 40	
33 34		35 36		37 40		ATTACHMENT SUBMITTED N 23 41		NPRD-4 FORM SUB. N 24 42		PRIME COMP. SUPPLIER N 25 43		COMPONENT MANUFACTURER X 9 9 9 26 44 47			

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

1 0 After investigation, the worker replaced the sample pump head assemblies
1 1 and diaphragms and cleaned the bypass check valves. The system was then
1 2 satisfactorily functionally tested per procedure and returned to service
1 3 on 2/28/83. "A" and "B" H2 O2 analyzers will be replaced in the future
1 4 with an improved model per Design Change Request 81-165.

							FACILITY STATUS								% POWER						OTHER STATUS						METHOD OF DISCOVERY						DISCOVERY DESCRIPTION													
							E						069						NA						B						Routine Surveillance															
							(28)						(29)												(30)						(31)						(32)									

ACTIVITY CONTENT
RELEASED OF RELEASE

1 6 33 34 NA

7 8 9 10 11 44

AMOUNT OF ACTIVITY (35)

NA

45 80

LOCATION OF RELEASE (36)

PERSONNEL EXPOSURES									
NUMBER			TYPE	DESCRIPTION					
1	7	0	0	0	37	Z	38	NA	39

PERSONNEL INJURIES		NUMBER		DESCRIPTION	
1	2	3	4	5	6
0	0	0	0	40	NA

LOSS OF OR DAMAGE TO FACILITY (43)
TYPE DESCRIPTION
1 9 Z (42) NA
8707710000 800810

8303210320 830310
PDR ADOCK 05000366
S PDR

PUBLICITY
ISSUED DESCRIPTION (45) NA

2 0 [N] (44) NA

NRC USE ONLY

NAME OF PREPARER S. B. Tipps

PHONE (912) 367-7851