

# LICENSEE EVENT REPORT

CONTROL BLOCK: 1

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 | MIDCC1 | 200-0000-00 | 341111 | 41111 | 5

CON'T  
 REPORT SOURCE: L605000315 | 7081583 | 8090883 | 9

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES 10

012 | WITH THE REACTOR COOLANT SYSTEM IN MODE 6, THE TRAIN A LOWER CONTAINMENT AIRBORNE  
013 | MONITOR (ERS-1300) WAS DECLARED INOPERABLE, DUE TO A FAILED CONDITION. THIS EVENT WAS  
014 | NON-CONSERVATIVE WITH RESPECT TO TECHNICAL SPECIFICATION 3.3.3.1 TABLE 3.3-6 ITEM 2.a  
015 | AND 3.9.9. THE ACTION REQUIREMENTS WERE MET. PUBLIC AND HEALTH AND SAFETY WERE NOT  
016 | AFFECTED. PREVIOUS OCCURRENCES INCLUDE: 50-315/83-038, 82-087.

017 |  
018 |

SYSTEM CODE: BB 11  
 CAUSE CODE: E 12  
 CAUSE SUBCODE: G 13  
 COMPONENT CODE: INSTRU 14  
 COMP SUBCODE: E 15  
 VALVE SUBCODE: Z 16  
 LER/RO REPORT NUMBER: 17  
 EVENT YEAR: 83  
 SEQUENTIAL REPORT NO.: 082  
 OCCURRENCE CODE: 03  
 REPORT TYPE: L  
 REVISION NO.: 0  
 ACTION TAKEN: X 18  
 FUTURE ACTION: Z 19  
 EFFECT ON PLANT: Z 20  
 SHUTDOWN METHOD: Z 21  
 HOURS: 0000  
 ATTACHMENT SUBMITTED: N 22  
 NPD-4 FORM SUB: Y 23  
 PRIME COMP. SUPPLIER: L 24  
 COMPONENT MANUFACTURER: E070

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS 27

110 | INVESTIGATION FOUND A PIECE OF METAL ON ONE OF THE CIRCUIT BOARD CARD CONNECTORS. ALL  
111 | CIRCUIT BOARDS WERE REMOVED AND CLEANED TO INSURE NO FOREIGN MATERIALS WERE PRESENT.  
112 | CIRCUIT BOARDS WERE REINSTALLED, MONITOR VERIFIED TO OPERATE CORRECTLY AND RETURNED TO  
113 | SERVICE.

114 |

FACILITY STATUS: H 28  
 % POWER: 000  
 OTHER STATUS: NA 29  
 METHOD OF DISCOVERY: B 30  
 DISCOVERY DESCRIPTION: OPERATOR OBSERVATION 31

ACTIVITY CONTENT: Z 32  
 RELEASED OF RELEASE: Z 33  
 AMOUNT OF ACTIVITY: NA 34  
 LOCATION OF RELEASE: NA 35

PERSONNEL EXPOSURES: 000  
 TYPE: Z 36  
 DESCRIPTION: NA 37

PERSONNEL INJURIES: 000  
 TYPE: NA 38  
 DESCRIPTION: NA 39

LOSS OF OR DAMAGE TO FACILITY: Z 40  
 TYPE: NA 41  
 DESCRIPTION: NA 42

PUBLICITY: N 43  
 ISSUED DESCRIPTION: NA 44

NAME OF PREPARER: R. A. PALMER  
 8309160220 830908  
 PDR ADOCK 05000315  
 S PDR

NRC USE ONLY

PHONE: 616-465-5901