

WASHINGTON STATE UNIVERSITY

PULLMAN, WASHINGTON 99164-1300

NUCLEAR RADIATION CENTER

November 15, 1983

Mr. Ross A. Scarano, Director
Division of Radiological Safety
and Safeguards Programs
U.S. Nuclear Regulatory Commission
Region V
1450 Maria Lane, Suite 210
Walnut Creek, California 94596

REGION V 11/16


1983 NOV 17 PM 2:32

RECEIVED
NRC

Dear Mr. Scarano:

In accordance with the agreement reached during a recent conference telephone call between members of your staff, myself, and Reactor Licensing, W.S.U. proposes to rectify the problem relating to the CAM calibration by requesting a change in the W.S.U. facility technical specifications. Specifically, a change to section 4.33 on Radiation Monitoring Systems Surveillance requirements to specify a "channel check" for the CAM rather than a full calibration is being requested. The change will more correctly describe the surveillance activities being carried out on the CAM system which licensing agrees are adequate.

Sincerely,



W. E. Wilson
Associate Director

WEW:efm

8312060321 831130
PDR ADOCK 05000027
Q PDR