

Arizona Public Service Company

P.O. BOX 21666 • PHOENIX, ARIZONA 85036

March 16, 1984
ANPP-29083-BSK/TRB

U. S. Nuclear Regulatory Commission
Region V
Creskide Oaks Office Park
1450 Maria Lane - Suite 210
Walnut Creek, CA 94596-5368

Attention: Mr. T. W. Bishop, Director
Division of Resident
Reactor Projects and Engineering Programs

Subject: Interim Report - DER 84-08
A 50.55(e) Potentially Reportable Deficiency Relating to ITT
Barton Model 763 Transmitters Do Not Meet Specifications.
File: 84-019-026; D.4.33.2


Reference: Telephone Conversation between R. Dobbs and T. Bradish on
February 29, 1984.

Dear Sir:

The NRC was notified of a potentially reportable deficiency in the referenced telephone conversation. At that time, it was estimated that a determination of reportability would be made within thirty (30) days.

Due to the extensive investigation and evaluation required, an Interim Report is attached. It is now expected that this information will be finalized by May 18, 1984, at which time a complete report will be submitted.

Very truly yours,



E. E. Van Brunt, Jr.
APS Vice President, Nuclear
ANPP Project Director

EEVB/TRB:db
Attachment

cc: See Page Two

8403300096 840316
PDR ADOCK 05000528
S PDR

1E-27 1/1

Mr. T. W. Bishop
DER 84-08
Page Two

cc: Richard DeYoung, Director
Office of Inspection and Enforcement
U. S. Nuclear Regulatory Commission
Washington, D. C. 20555

T. G. Woods, Jr.
W. E. Ide
D. B. Fasnacht
A. C. Rogers
B. S. Kaplan
L. A. Souza
J. Vorees
J. R. Bynum
P. P. Klute
A. C. Gehr
W. J. Stubblefield
W. G. Bingham
R. L. Patterson
R. W. Welcher
H. D. Foster
D. R. Hawkinson
L. E. Vorderbrueggen
G. A. Fiorelli
S. R. Frost
J. Self
D. Canady

Records Center
Institute of Nuclear Power Operations
1100 Circle 75 Parkway, Suite 1500
Atlanta, GA 30339

INTERIM REPORT - DER 84-08
POTENTIAL REPORTABLE DEFICIENCY
ARIZONA PUBLIC SERVICE COMPANY (APS)
PVNGS UNIT 1, 2, 3

I. Potential Problem

Per the referenced C-E letter the elevated/suppressed range 763 transmitters have errors of -2 to -3% of calibrated span. The affected Bechtel tag numbers are:

1,2&3JRCEPT100X	1,2&3RCCPT101C	1,2&3SIAPT333
1,2&3JRCEPT100Y	1,2&3RCCPT101D	1,2&3SIAPT343
1,2&3JRCAPT101A	1,2&3S1BPT313	2&3RCAPT103
1,2&3JRCBPT101	1,2&3S1BPT323	2&3RCBPT104

II. Approach To and Status of Proposed Resolution

Bechtel Engineering is currently corresponding with C-E to determine the impact of design modifications to existing PVNGS transmitters.

III. Projected Completion of Corrective Action and Submittal of the Final Report

Evaluation of this condition and submittal of the Final Report is Forecast to be completed by May 18, 1984