

From: [Cender, Laura](#)
To: [Jim Botti](#)
Subject: Pending Renewal of NRC License No. 21-32196-01 - Beaumont Heart and Vascular Center
Date: Thursday, April 16, 2020 9:36:00 AM

Hello Jim,

I am currently processing the license renewal request for Beaumont Heart & Vascular Center, NRC License No. 21-32196-01, and I have a few of questions about the application submitted.

1. Section 8.5 and 8.6 of the submitted application request authorization for both 10 CFR 35.100 and 10 CFR 35.200. Currently the license only lists 35.200 and the materials and uses requested for authorized users in Section 8.7.2 is limited to 10 CFR 35.200 as well. Please clarify if 10 CFR 35.100 uses of material are requested, and if so please indicate the AU(s) who would be authorized for 35.100 materials and uses.
2. Item 8.9.1 of the submitted application states that two facility diagrams were enclosed, but only one facility diagram was received. If a second facility diagram was intended to have been included, please submit a copy with your response.
3. Please indicate on the facility diagram which doors are access controlled (i.e. locked).

Please ensure that your response is dated and is signed by either the RSO, Gregory Pitt, CNMT, or a senior manager for the licensee.

Please feel free to contact me at 630-829-9712 or send me an email if you have any questions.

Thank you,
Laura

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712