



# MILLER ENGINEERING & TESTING, INC.

GEOTECHNICAL / SOIL BEHAVIORS / ENVIRONMENTAL / CONCRETE / STEEL / ROOFING / ASPHALT INSPECTION

**Mail all correspondence to:**

100 SHEFFIELD ROAD • P.O. BOX 4776 • MANCHESTER, NEW HAMPSHIRE 03108 • TELEPHONE (603) 668-6016 • FAX (603) 668-8641

August 26, 1992

Mr. Francis M. Costello, Chief  
Industrial Application Section  
Division of Radiation Safety & Safeguards  
U.S. NUCLEAR REGULATORY COMMISSION  
Region 1  
475 Allendale Road  
King of Prussia, PA 29406-1415

Re: Routine Inspection No. 030-22208/92-001  
Docket No. 030-22208  
License No. 28-23457-01

Dear Mr. Costello:

As promised in my letter to you dated August 17, 1992, I am enclosing a copy of a Certificate of Leak Test for Nuclear Moisture/Density Gauge, Serial No. 5430, dated July 13, 1992.

Thank you for your attention to this matter.

Very truly yours,

MILLER ENGINEERING & TESTING, INC.

Arthur W. Rose, P.E.  
Senior Vice President

AWR:paz

9212210161 921203  
PDR ADDOCK 03022208  
C PDR

AUG 28 1992

# CPN

a division of Brainard Kilman

2830 Howe Road  
Martinez, California 94553  
Telephone (510) 228-9770  
FAX (510) 228-3183

## CERTIFICATE OF LEAK TEST

RECEIVED AUG 24 1992

MILLER ENGINEERING & TESTING, INC.  
ATTN: Mr. Arthur Rose  
P.O. Box 4776  
Manchester,

NH 03108-0000

Customer Number:  
Leak Test Date : July 13, 1992  
Test Number : 3888  
RDC # : 143235

This is to certify that the leak test on the indicated source(s) had been counted on July 29, 1992 and the results shown accurately represent the level of removable contamination.

Model : MC-2      Serial Number : M24055430

Source	Activity	* Removable Contamination +or- Counter Error
-----	-----	-----
Am241/Be	50mCi	<0.001 nCi
Cs137	10mCi	<0.01 nCi

\* (Note: Removable contamination equal to or greater than 5.0 nanoCuries constitutes a leaking source.)

Reviewed by :  
Sean Reilly  
Assistant RSO



BRAINARD-KILMAN

*A Company Committed to Quality and Reliability.*