

APR 30 1985

DOCKET NO. 030-22126
CONTROL NO. 03553

Eastern Aquanalysis
ATTN: Michael Hagar
Chemist
Box 700
Brooklyn, CT 06234

Gentlemen:

This is in reference to your application dated March 14, 1985, for a Byproduct Material License. In order to continue our review, we need the following additional information:

1. Please submit the detector cell model numbers you will use for the Tracor Model 560 gas chromatograph.
2. Please submit your procedures for cleaning the detector cells and for removal and exchange of detector cell foils. If you follow the manufacturer's recommended procedures then please provide a copy of the procedures. If detector cell are returned to your supplier for cleaning of the detector cells or removal and exchange of the foils, then please specify this.

We will continue our review of your application upon receipt of the above information. Please reply in duplicate, referencing Control No. 03123.

Sincerely,

Original Signed By:

John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

8506170252 850531
REG1 LIC30
06-20823-01 PDR

RI:DRSS

Varela

MA
4/29/85

RI:DRSS
Glenn



4/29/85

"OFFICIAL RECORD COPY"

ML10