

ENCLOSURE

NOTICE OF VIOLATION

Department of Navy  
Washington, D.C.

Docket No. 030-29462  
License No. 45-23645-01NA  
Permit No. 12-00211-11NP

During an NRC inspection conducted on November 23, 1992, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, the violation is listed below:

10 CFR 35.51(c) requires, in part, that a licensee check each survey instrument for proper operation with the dedicated check source each day of use.

Contrary to the above, between March 15, 1992 and November 23, 1992, the licensee routinely did not check its survey meter with a dedicated check source on days when the instrument was used.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, the Department of the Navy is hereby required to submit a written statement or explanation to the U.S. Nuclear Regulatory Commission, ATTN: Document Control Desk, Washington, D.C. 20555 with a copy to the Regional Administrator, Region II, within 30 days of the date of the letter transmitting this Notice of Violation (Notice). This reply should be clearly marked as a "Reply to a Notice of Violation" and should include: (1) the reason for the violation, or, if contested, the basis for disputing the violation, (2) the corrective steps that have been taken and the results achieved, (3) the corrective steps that will be taken to avoid further violations, and (4) the date when full compliance will be achieved. If an adequate reply is not received within the time specified in this Notice, an order or demand for information may be issued as to why the license should not be modified, suspended, or revoked, or why such other action as may be proper should not be taken. Where good cause is shown, consideration will be given to extending the response time.

Dated at Atlanta, Georgia  
this *4th* day of *December*, 1992

9212300097 921204  
PDR ADOCK 03029462  
C PDR