

United States Nuclear Regulatory Commission
Region III
799 Roosevelt Rd.
Glen Ellyn, IL 60137
Materials Licensing Section
Attn: Evelyn Matson

May 20, 1985

Control Number: 17882

License Number # 12-15842-01

Dear Evelyn;

Enclosed is our license for your review. Everything from what I can see is pretty straight forward. I had pulled appendices from the guide (#10.8) and they have been copied inadvertently and put in with these copies. So please excuse the "thickness" of this application for renewal.

I would like to address the items mentioned in your letter dated April 9, as requested:

1&2: Procedures for handling therapeutic quantities of iodine-131:
Past experience has shown that there has been no problems with liquid I131. However, we will in the future, use the capsule form of I131, or liquid vials, which shall not be opened at anytime and administered thru a needle sipper.

3: Our method for determining that the ventilation system in Xenon-133 use areas meets our specification is done semi-annually by Certified Test & Balance Co. in Chicago. We maintain records of results and adjust air-flow rates if needed.

I would appreciate hearing from you soon.

Sincerely,

Judy Meyers

Judy Meyers, Technical Director

*Wilbur Hospital
975 N 5th Ave*

St Charles, IL 60174

RECEIVED

MAY 23 1985

REGION III

8506190029 850604
REG3 LIC30
12-15842-01 PDR