

25 APR 1988

License No. 29-03047-01  
Docket No. 30-02465  
Control No. 108569

Beth Israel Hospital  
ATTN: Jessie Z. Trivino, M.S.  
Radiation Safety Officer  
70 Parker Avenue  
Passaic, NJ 07055

Gentlemen:

This is in reference to your request in a letter dated February 4, 1988 to Amend License No. 29-03047-01. In order to continue our review, we need the following additional information:

35.70 of 10 CFR 35 requires that licensees maintain records of surveys for removable contamination in units of disintegrations per minute. Please revise the weekly survey form submitted in your amendment request to express the results of the wipe surveys in units of disintegrations per minute.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,

Original Signed By:  
John E. Glenn

John E. Glenn, Ph.D., Chief  
Nuclear Materials Safety Section A  
Division of Radiation Safety  
and Safeguards

RI:DRSS  
Tripp/mk

RI:DRSS  
Glenn

04/21/88

04/24/88

OFFICIAL RECORD COPY

ML 334 TRIPP - 0001.0.0  
04/20/88

9301060379 920520  
PDR FOIA  
STOLL 92-58

PDR

ML10

D17