



## CORPORATION

338 DELAWANNA AVE.

• CLIFTON, N.J. 07014 •

PHONE: 201-773-9400 •

TLX: 133559 XID CLIF

MS 16  
PS

May 20, 1985

Nuclear Regulatory Commission

Attn: Mr. John E. Glenn, Ph. D.

RE: LICENSE NO. 29-13064-02 CONTROL NO. 17428

Dear Sir,

Enclosed please find a sample of our Certificate of Calibration we submit with each calibration completed, as requested by the NRC.

Sincerely,

Eileen Radowitz  
Office Manager

8506190154 850603  
REG1 LIC30  
29-13064-02 PDR

RECEIVED-REG  
MAY 23 PM 2:37

17428  
"OFFICIAL RECORD COPY"

ML10

MAY 23 1985



# CORPORATION

P. O. BOX 1015

• 338 DELAWANNA AVE. •

CLIFTON, N.J. 07014 •

PHONE: 201 - 773-9400

## CERTIFICATE OF CALIBRATION

NRC license # 29-13064-02

Client:

*Warrington Div.  
McGraw Edison*

Date:

*9/6/84*

Manufacturer:

*Victoreen*

Electronic Test:

*OK*

Model Number:

*592B*

Serial Number:

*2651*

<u>RANGE</u>	<u>CALIBRATION POINT</u>	<u>ACTUAL READING</u>
X1	2.0 Mr/Hr	<i>2.0</i>
X1	8.0 Mr/Hr	<i>8.0</i>
X10	20 Mr/Hr	<i>20</i>
X10	80 Mr/Hr	<i>80</i>
X100	200 Mr/Hr	<i>200</i>
X100	800 Mr/Hr	<i>800</i>

Repairs/Parts

*6 - 225 Volt batteries*

This instrument should be re-calibrated no later than 90 days from last date of calibration.

The calibration and certification is in accordance with the Nuclear Regulatory Commission requirements of 10 CFR Part 34 Regulations.

Calibration Source #S 318, CS-137, TECH/OPS

Radiological Services Division

Instrument Repair and Calibration

Calibrated by: \_\_\_\_\_

*[Signature]*

5/20/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

11:05

☐ A.M.  
☐ P.M.

☐ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

J. A. Jaustra

OFFICE/ADDRESS

Region I

PHONE NUMBER

EXTENSION

5257

PERSON CALLED

Marty Cogan

OFFICE/ADDRESS

XID Corp.

PHONE NUMBER

EXTENSION

201  
773-9400

CONVERSATION

SUBJECT

SUMMARY

we need verification whether a certificate  
is given to those who  
receive XID Services.  
(Calibration Certificate)

they will forward a copy of certificate.

REFERRED TO:

ACTION REQUESTED

☐ ADVISE ME OF  
ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE

"OFFICIAL RECORD COPY"

MLT