

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

24-18628-01

Docket or Reference number

030-13966

Amendment No. 04

North Kansas City Memorial Hospital
2800 Hospital Drive
North Kansas City, MO 64116

In accordance with letter dated April 24, 1985, License Number 24-18628-02 is amended as follows:

Items 6., 7., 8., and 9. are amended to add:

6. Byproduct, source,
and/or special nuclear
material

7. Chemical and/or
physical form

8. Maximum amount that
licensee may possess
at any one time
under this license

I. Americium-241

I. Sealed source
(Amersham Model
AMC.24)

I. 15 millicuries

9. Authorized Use

I. For use in a Siemens Gammasonics Model 035-42300C Dual Isotope Motion Correction Point Source Holder.

Condition 20. is amended to read:

20. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated May 25, 1984; letters dated March 5, 1985 and April 24, 1985; and Model ALARA Program as contained in Appendix O of Regulatory Guide 10.8, October 1980. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

For the U.S. Nuclear Regulatory Commission

Original Signed

By B.J. Holt

Materials Licensing Section, Region III

Date May 17, 1985

8506100770 850517
REG LIC30
24-18628-01 PDR

COPY
ML30/BK MAY 17 1985