

OCT 29 1984

License No. 29-13064-02
Docket No. 030-14992
Control No. 17428

XID Corp.
ATTN: Marty Cogan
Vice President
P. O. Box 908
338 Delaware Avenue
Clifton, New Jersey 07014

Gentlemen:

This is in reference to your application dated April 11, 1984, to renew License Number 29-13064-02. In order to continue our review, we need the following additional information:

1. Please submit your detailed procedures for calibrating survey meters. Include in this description the method used in positioning the source and meter to be calibrated, distances used for each reading being obtained and method used in determining the accuracy of each reading.
2. Please specify the activity of your Cs-137 sealed source used in Tech/Ops model number 77302, and submit your calculations which were used in correcting for decay of this sealed source.
3. Please submit a diagram of your facility, please indicate the type, dimensions, positions and thickness of shielding used, storage location of sources and identify adjacent areas across the wall from use and storage locations and show that adequate steps have been taken to assure that radiation levels in unrestricted areas do not exceed limits specified in 10 CFR 20.105 (enclosed).
4. Please submit a detailed description of your training program. Include in this description the topics covered, the duration of training and the method used to determine the trainees competency.

3506190165 850603
REG1 LIC30
29-13064-02 PDR

OFFICIAL RECORD COPY

366JOUSTR10/12/84 - 0001.0.0
10/23/84

ML10

XID Corp.

2

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 17428.

Sincerely,

Original Signed By:
John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Section B
Division of Engineering and
Technical Programs

Enclosure: 10 CFR Part 20

RI:DETP ³⁰⁶
Joustra/cop
10/25/84

RI:DETP
Glenn
10/29/84

OFFICIAL RECORD COPY

366JOUSTRA10/12/84 - 0002.0.0
10/23/84