

APR 05 1985

License No. 37-20778-01
Docket No. 030-22064
Control No. 03506

James A. Nicotero, M.D.
4401 Penn Avenue
Suite 1050
Pittsburgh, Pennsylvania 15224

Dear Dr. Nicotero:

This is in reference to your request in a letter dated March 5, 1985 to amend License No. 37-20778-01. In order to continue our review, we need the following additional information:

1. Please supply a letter from the administrator/manager or each site authorizing you to use the densitometer at their facility. Also, have the administrator/manager indicate where in the facility you are allowed to use the densitometer. If a single room is designated, please submit a diagram of the room and adjacent rooms.
2. Please confirm that only an authorized physician user will transport the densitometer to the various temporary jobsites. Note that 10 CFR Part 71.9 allows only physicians operating under the license issued in accordance with 10 CFR Part 35 to transport licensed material for the practice of medicine. The alternative to 10 CFR Part 71.9 is to use a common or private carrier to transport the licensed material in accordance with 10 CFR Part 71.5.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 03506.

Sincerely,

Original Signed By:
Jenny M. Johansen

8506190252 850524
REG1 LIC30
37-20778-01 PDR

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

Enclosures:
10 CFR Parts 35 and 71

RI:DRSS
Foster/djh
3/18/85

RI:DRSS
Glenn
4/5/85

0001.0.0

OFFICIAL RECORD COPY

JAMES A NICOTERO MD/D/LTR -

03/25/85

ML10