

DEPARTMENT OF THE AIR FORCE
HEADQUARTERS AIR FORCE SPECIAL WEAPONS CENTER (AFSWC)
KIRTLAND AIR FORCE BASE, NEW MEXICO 87117



REPLY TO
ATTN OF: SWD

8 April 1968

SUBJECT: AEC Compliance Inspection (AEC Byproduct Material License No. 30-03110-02
and Special Nuclear Material License No. SNM-597) 70-647

TO: United States Atomic Energy Commission
Division of Compliance, Region IV
10395 West Colfax, Room 200
Denver, Colorado 80215

ATTN: Mr. Donald I. Walker
Director

1. Reference your letter, 27 March 1968, pertaining to subject inspection.
2. Now that more realistic cycling periods for leak testing sources have been established, all sources requiring a leak test will be tested at least each six months. This discrepancy had been corrected prior to the inspection inasmuch as all sources were in compliance within the past six months.
3. Leak test records, effective with subsequent leak test data, will be recorded in units of microcuries.

W. R. MORTON, Colonel, USAF
Vice Commander

2301 01 9-68

005272

8507300396 850530
PDR FOIA
BURR85-229 PDR