

official

DEC 18 1992

Docket No. 030-32887
License No. 45-25216-01

Community Radiology of Virginia, Inc.
ATTN: Michael Shahan, M.D.
Vice-President
2000 Leatherwood Lane
Bluefield, VA 24605

Gentlemen:

SUBJECT: NOTICE OF VIOLATION
(NRC INSPECTION REPORT NO. 45-25216-01/92-01)

This refers to the inspection conducted by Mr. W. Loo of this office on December 8, 1992. The inspection included a review of activities authorized for your Community Radiology of Virginia, Inc. facility. At the conclusion of the inspection, the findings were discussed with you.

The inspection was an examination of activities conducted under your license with respect to radiation safety and compliance with NRC regulations and the conditions of your license. It included selective examinations of procedures and representative records, interviews with personnel and direct observations by the inspector.

Based on the results of this inspection, certain of your activities appeared to be in violation of NRC Requirements, as specified in the enclosed Notice of Violation. In addition, the inspector identified activities that violated NRC requirements that will not be subject to enforcement action because the licensee's efforts in identifying and/or correcting the violations meet the criteria specified in Section VII.B of the Enforcement Policy. Those non-cited violations included the failure to maintain records of dose calibrator tests, survey instrument calibrations, weekly removable contamination surveys and performance tests for transportation packages and failure to post required documents.

You are required to respond to this letter and should follow the instructions specified in the enclosed Notice when preparing your response. In your response you should document the specific actions taken and any additional actions you plan to prevent recurrence. After reviewing your response to this Notice, including your proposed corrective actions and the results of future inspections, the NRC will determine whether further NRC enforcement action is necessary to ensure compliance with NRC regulatory requirements.

In accordance with 10 CFR 2.790 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be placed in the NRC's Public Document Room.

9301080018 921218
PDR ADOCK 03032887
C PDR

IE07

Community Radiology of Virginia, Inc. 2

The responses directed by this letter and the enclosed Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, Pub. L. No. 96-511.

Should you have any questions concerning this letter, please contact us.

Sincerely,

*original signed by
J. Potter for*

Douglas M. Collins, Chief
Nuclear Materials Safety
and Safeguards Branch
Division of Radiation Safety
and Safeguards

Enclosure:
Notice of Violation

cc w/encl:
State of Virginia

bcc w/encl:
Document Control Desk

R11:DRSS

WLoo (WTL)
12/15/92

R11:DRSS

Mooney
12/18/92