

AUG 26 1990

License No. 20-00742-18
Docket No. 030-09062
Control No. 112380

Beth Israel Hospital
ATTN: M. Rosemary Kennealy
Radiation Safety Officer
330 Brookline Avenue
Boston, Massachusetts 02215

Gentlemen:

Please find enclosed an amendment to your NRC Material License.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the Region I Material Licensing Section, (215) 337-5093, so that we can provide appropriate corrections and answers.

Please be advised that you must conduct your program involving licensed radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, please note the items in the enclosed, "Requirements for Materials Licensees."

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, the NRC expects licensees to pay meticulous attention to detail and to achieve the high standard of compliance which the NRC expects of its licensees.

You will be periodically inspected by NRC. A fee may be charged for inspections in accordance with 10 CFR Part 170. Failure to conduct your program safely and in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in prompt and vigorous enforcement action against you. This could include issuance of a notice of violation, or in case of serious violations, an imposition of a civil penalty or an order suspending, modifying or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

9301110046 920520
PDR FOIA
STOLL92-58 PDR

OFFICIAL RECORD COPY

ML 20-00742-18/LTR - 0001.0.0
06/26/90

9301110046 3PP ML 10

D/80

We wish you success in operating a safe and effective licensed program.

Sincerely,

Original Signed By:
Jenny Johansen

for Judith A. Joustra
Health Physicist
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

Enclosures:

1. Amendment No. 14
2. Requirements for Materials Licensees
3. Requirements for Medical Licensees
4. NRC Forms 3 and 313
5. 10 CFR Parts 2, 19, 20, 35, and 170
6. Regulatory Guides 10.8

DRSS:RI
Darden/cmm
J. H. Darden
7/8/13/90

DRSS:RI
Joustra *Joustra*
08/22/90
for

OFFICIAL RECORD COPY

ML 20-00742-18/LTR - 0002.0.0
06/26/90

4/20/91

PAGE 1

U.S. NUCLEAR REGULATORY COMMISSION
REGION I
OPEN ITEMS TRACKING SYSTEM

DOCKET NUMBER: 30-09062

REPORT NUMBER	STATUS	DATE OPEN	DATE CLOSED	REVIEWER NAME	CLOSING OPTION	REFER
1 87-01	CLOSED	01/02/87	01/02/87	MURTI	VIOL	87-01
ITEM: CLEAR - 591						
2 85-01	CLOSED	02/01/85	01/02/87	JOHANGEN	VIOL	87-01
ITEM: APPROPRIATE PROTECTION MEASURES WHEN HANDLING RAM WERE NOT USED. (TONGS AND PLEXIGLASS)						
3 87-198	CLOSED	11/20/87	05/24/89	BETSY ULLRICH	LER	89-001
ITEM: TWO (2) DIAGNOSTIC MISADMINISTRATIONS						
4 81-001	OPEN	09/04/91	/ /	COURTEMANCHE	VIOL	
ITEM: 20.202 RAD MONITOR EQUIP NOT PROVIDED TO PERSONNEL HANDLING BRACHY DEVICES (JOINT COMMISSION)						

Δ/82