



E. I. DU PONT DE NEMOURS & COMPANY  
INCORPORATED

GLENOLDEN LABORATORY  
GLENOLDEN, PENNSYLVANIA 19036

CENTRAL RESEARCH & DEVELOPMENT DEPARTMENT

May 21, 1985

U.S. Nuclear Regulatory Commission  
Region I  
631 Park Avenue  
King of Prussia, PA 19406

Attention: J.D. Kinneman

RE: NRC Inspection No. 030-19346/85-01  
DuPont - Glenolden Laboratory  
License No. 07-00455-37


Dear Sir:

In response to your notice of violation stemming from the subject inspection conducted on January 29-30, we concur with findings of your inspector, Mr. R.H. Ladun.

The laboratory chair which he found to be contaminated was decontaminated immediately following the inspection and the rest of the laboratory was thoroughly surveyed for additional contamination, including laboratory coats and personal clothing worn by possible users of that particular work area. No other contamination was found either by survey meter or by wipe testing.

The importance of conducting surveys during and after work with millicurie quantities of P-32 has been communicated to all users of isotopes.

Sincerely,

  
George W. Moncrief  
Radiation Safety Officer

8506200468 850607  
REQ1 LIC30  
07-00455-37 PDR