

RADIATION CALIBRATION LABORATORY

1177 McCarter Highway • Newark, New Jersey 07104 • (201) 482-4809 • NRC Lic. No. 29 20880 01

December 10, 1992

Mr. Francis M. Costello, Chief - Industrial Applications Section
Division of Radiation Safety and Safeguards
Nuclear Regulatory Commission, Region 1
475 Allendale Road
King of Prussia, PA 19406-1415

Reference: Docket No. 030-28793
License No. 29-20880-01

Subject: Routine Inspection No. 92-001

Dear Mr. Costello:

In compliance with your letter dated 25 November 1992 and Appendix "A", Notice of Violation appended thereto, please be advised as follows:

1] The initial corrective actions which have been taken and the results achieved are that a physical inventory was performed on 3 November 1992. The results were recorded on the Radiation Calibration Laboratory "Radioactive Material Control and Leakage Test Record Form". The signature of the person performing the inventory was recorded.

The results of this inventory indicated that it was identical to that performed on 11 May 1990.

2] The corrective steps taken were to expand the "Radioactive Material Inventory Control and Leakage Test Record" to be able to record the physical inventories of the sealed calibration sources for a five and one half year time period on a single card.

In order to avoid further violations, a yearly calendar was marked with appropriate dates to alert the management to take the physical inventory.


3] The date when full compliance was achieved was 3 November 1992, the day following the routine safety inspection conducted by Mr. Richard Gibson, Jr. of your office.

On 12 November 1992, Mr. Michael T. Gasparik of Radiation Calibration Laboratory, called Mr. Gibson at 215-337-5071, to return his call of 10 November 1992 and to apprise him of the action taken by this NRC licensee. Unfortunately, Mr. Richard Gibson was not available and we were unable to convey the above information to your office.

Your understanding and cooperation is appreciated by us. If the above response does not comply with NRC regulations, please advise the undersigned.

Sincerely,

Radiation Calibration Laboratory


Michael T. Gasparik
President

9301110139 921223
PDR ADOCK 03028793
C PDR

DEC 16 1992