



May 29, 1981

Mr. Paul Guinn, Mail Stop 39655  
Materials Licensing Branch  
U.S. Nuclear Regulatory Commission  
Washington, D.C. 20555

Reference: Specific License Devices

Dear Mr. Guinn,

Enclosed is a list of devices that Ohmart distributes to Specific Licensees. This list includes obsolete devices as well as devices that we are currently manufacturing.

We certainly do approve of the method you proposed to distribute approved devices to Specific Licensees. If this method is adopted, we estimate that we would save about one hundred labor hours per month. More importantly, though, would be the labor savings at the NRC and the consequent reduced turn-around time for licensing. Frankly, the delay in obtaining a license or license amendment is one of our major problems in scheduling production.

We hope that the plan is adopted.

Sincerely,

H. L. Cook  
Vice President

HLC:ra

8506220258 850305  
PDR FOIA  
EDMUNDS85-31 PDR

See 6/24/81 memo  
attached

No. deb. combined  
w/02977  
08097

44

# THE OHMART CORPORATION

## Devices Distributed to Specific Licensees

<u>Device Model Numbers</u>	<u>Source Model Numbers</u>	<u>Isotopes</u>	<u>Maximum Activity Per Source (Millicuries)</u>	<u>Wipe Test Interval</u>
AS-2	A-2102	Cs-137	100	3 yrs.
ASR-2	A-2102 or A-2104	Cesium 137	4500	3 yrs.
ASR-3	A-2102 or A-2104	Cesium 137	4000	3 yrs.
ASR-4	A-2102	Cesium 137	2700	3 yrs.
BAL	A-33766	Krypton 85	75	Not Rec
	A-4829	Krypton 85	1200	Not Rec
	A-36058	Krypton 85	1500	Not Rec
	A-4830	Strontium 90	50	6 Mos.
	A-5799	Americium 241	1000	6 Mos.
	A-22439	Iron 55	150	6 Mos.
Beta Art	A-35950	Krypton 85	75	Not Rec
Series 6000	A-36058	Krypton 85	200	Not Rec
(with BAL source holder)	A-22439	Iron 55	150	6 Mos.
BG-1G	A-4829 or A-4834	Krypton 85	600	Not Rec
BG-2G	A-5840 or A-4830	Strontium 90	300	6 Mos.
BG-2/9G	B-14315	Strontium 90	100	6 Mos.
	B-14315	Krypton 85	1200	Not Rec
BG-3G	A-4832, A-4831 or A-5800	Strontium 90	1000	6 Mos.
BG-4G	B-6815	Krypton 85	1000	Not Rec
BG-9G	A-4829 or A-4834	Krypton 85	1200	Not Rec
BG-11G	B-6815	Krypton 85	2000	Not Rec
BGCD-2/9G	B-14315	Strontium 90	100	6 Mos.
	B-14315	Krypton 85	1200	Not Rec
BGCD-9G	A-4829	Krypton 85	1200	Not Rec
BGO-1G	A-4829 or A-4834	Krypton 85	600	Not Rec
BGO-2G	A-5840 or A-4830	Strontium 90	300	6 Mos.
BGO-3G	A-4832, A-4831, or A-5800	Strontium 90	1000	6 Mos.
BGO-4G	B-6815	Krypton 85	1000	Not Rec
BGO-9G	A-4829 or A-4834	Krypton 85	1200	Not Rec
BGO-11G	B-6815	Krypton 85	2000	Not Rec
BGOC-1G	A-4829 or A-4834	Krypton 85	600	Not Rec
BGOC-2G	A-5840 or A-4830	Strontium 90	300	6 Mos.
BGOC-3G	A-4832, A-4831 or A-5800	Strontium 90	1000	6 Mos.
BGOC-4G	B-6815	Krypton 85	1000	Not Rec
BGOC-9G	A-4829 or A-4834	Krypton 85	1200	Not Rec
BGOC-11G	B-6815	Krypton 85	2000	Not Rec
BGOM-1G	A-4829 or A-4934	Krypton 85	600	Not Rec
BGOM-2G	A-5840 or A-4830	Strontium 90	300	6 Mos.
BGOM-3G	A-4832, A-4831 or A-5800	Strontium 90	1000	6 Mos.
BGOM-4G	B-6815	Krypton 85	1000	Not Rec

<u>Device Model Numbers</u>	<u>Source Model Numbers</u>	<u>Isotopes</u>	<u>Maximum Activity Per Source (Millicuries)</u>	<u>Wipe Test Interval</u>
BGOM-9G	A-4829 or A-4834	Krypton 85	1200	Not Rec
BGOM-11G	B-6815	Krypton 85	2000	Not Rec
BGOU-9G	A-4829	Krypton 85	1200	Not Rec
BGTL-2/9G	B-14315	Strontium 90	100	6 Mos.
	B-14315	Krypton 85	1200	Not Rec
BGTL-9G	A-4829	Krypton 85	1200	Not Rec
BWGL	A-5771	Cesium 137	100	3 yrs.
CC	A-5776	Cesium 137	3	3 yrs.
CL-10	A-2102	Cesium 137	3000	3 yrs.
CL-12	A-2102 or A-2104	Cesium 137	3750	3 yrs.
CL-14	A-2102 or A-2104	Cesium 137	4500	3 yrs.
CL-16	A-2102 or A-2104	Cesium 137	5250	3 yrs.
CL-18	A-2102 or A-2104	Cesium 137	7500	3 yrs.
CL-20	A-2102 or A-2104	Cesium 137	9400	3 yrs.
CL-30	A-2102 or A-2104	Cesium 137	5000	3 yrs.
Coke Drum Level	A-35210	Cesium 137	500	3 yrs.
	B-38239	Cobalt 60	25	3 yrs.
CP-2	A-2102 or A-2104	Cesium 137	4000	3 yrs.
CP-3	A-2102 or A-2104	Cesium 137	4800	3 yrs.
CP-4	A-2102 or A-2104	Cesium 137	5000	3 yrs.
CP-5	A-2102 or A-2104	Cesium 137	5000	3 yrs.
CP-6	A-2102 or A-2104	Cesium 137	6000	3 yrs.
CP-8	A-2102 or A-2104	Cesium 137	6000	3 yrs.
CS-2	A-2102 or A-2104	Cesium 137	4500	3 yrs.
CS-3	A-2102 or A-2104	Cesium 137	4000	3 yrs.
CS-4	A-2102	Cesium 137	2700	3 yrs.
ED-2	A-2102	Cesium 137	300	3 yrs.
ED-3	A-2102	Cesium 137	500	3 yrs.
ED-4	A-2102	Cesium 137	750	3 yrs.
ED-5	A-2102	Cesium 137	1000	3 yrs.
ED-6	A-2102	Cesium 137	1500	3 yrs.
ED-8	A-2102	Cesium 137	2000	3 yrs.
ED-10, ED-12	A-2102 or A-2104	Cesium 137	3000	3 yrs.
ED-14	A-2102 or A-2104	Cesium 137	4000	3 yrs.
ED-16 ED-18 ED-20	A-2102 or A-2104	Cesium 137	5000	3 yrs.
ES-2	A-2102 or A-2104	Cesium 137	4500	3 yrs.
ES-3	A-2102 or A-2104	Cesium 137	4000	3 yrs.
ES-4	A-2102 or A-2104	Cesium 137	2700	3 yrs.
GRS-1	A-2102	Cs-137	1000	3 yrs.
GRS-2	A-2102	Cs-137	2000	3 yrs.
HM-8	A-2102	Cs-137	5000	3 yrs.
HM-10	A-2102	Cs-137	20000	3 yrs.

<u>Device Model Numbers</u>	<u>Source Model Numbers</u>	<u>Isotopes</u>	<u>Maximum Activity Per Source (Millicuries)</u>	<u>Wipe Test Interval</u>
LASR-2	A-2102 or A-2104	Cesium 137	4500	3 yrs.
LASR-3	A-2102 or A-2104	Cesium 137	4000	3 yrs.
LASR-4	A-2102	Cesium 137	2700	3 yrs.
LBG-1G	A-4829 or A-4834	Krypton 85	600	Not Req
LBG-2G	A-5840 or A-4830	Strontium 90	300	6 Mos.
LBG-3G	A-4832, A-4831 or A-5800	Strontium 90	1000	6 Mos.
LBG-4G	B-6815	Krypton 85	1000	Not Req
LBG-9G	A-4829 or A-4834	Krypton 85	1200	Not Req
LBG-11G	B-6815	Krypton 85	2000	Not Req
LR-10 or LRV-10	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LR-12 or LRV-12	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LR-14 or LRV-14	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LR-16 or LRV-16	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LR-20 or LRV-20	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LR-24 or LRV-24	A-2102 or A-2104	Cesium 137	9500	3 yrs.
LSDG	A-2102 or A-2104	Cesium 137	5400	3 yrs.
MDDG	See AS-2			
MDL	See AS-2			
NMG-1 for conveyors, pipes or chutes				6 Mos.
	B-18248	Am/Be or Pu/Li and Cs-137 or Co-60	1-2000  1- 2 gram  2-50  2-5	
	B-18489	Am/Be and Cs-137 or Co-60	2-2000  3-25  3-5	
NMG-2 for conveyors				6 Mos.
	A-33131 A-2102	Cf-252 Cs-137	2.5 microgram 50	
PG-2 through PG-4 or PGV-2 through PGV-4	A-2102	Cesium 137	1500	3 yrs.
PG-5 or PGV-5	A-2102	Cesium 137	1050	3 yrs.

<u>Device Model Numbers</u>	<u>Source Model Numbers</u>	<u>Isotopes</u>	<u>Maximum Activity Per Source (Millicuries)</u>	<u>Wipe Test Interval</u>
PG-6 or PG-8 or PGV-6 or PGV-8	A-2102	Cesium 137	1000	3 yrs.
PR-1G	A-4829 or A-4834	Krypton 85	600	Not Req
PR-9G	A-4829 or A-4834	Krypton 85	1200	Not Req
Ray-Weigh Scale - See SHRM Series				
RTSN-3	A-5776	Cesium 137	3	3 yrs.
SDG	A-2102 or A-2104	Cesium 137	5400	3 yrs.
SH-100	A-2102	Cesium 137	100	3 yrs.
SHD	A-2102	Cs-137	500	3 yrs.
SHDP or SHSP	A-33361	Cesium 137	200	3 yrs.
SHGS-2	A-2102	Cs-137	500	3 yrs.
SHLG-1	A-2102	Cs-137	2400	3 yrs.
	A-2100	Co-60	30	3 yrs.
SHLG-2	A-2104	Cs-137	9000	3 yrs.
	A-25159	Cs-137	4000	3 yrs.
	A-2100	Co-60	75	3 yrs.
SHLG-3	A-25159	Cs-137	4000	3 yrs.
	A-2104	Cs-137	9000	3 yrs.
	A-2100	Co-60	750	3 yrs.
SHLG-3A	A-22593	Cs-137	10,000	3 yrs.
SHLM-B & SHLM-C				
-1	A-2102	Cs-137	100	3 yrs.
	A-5771	Cs-137	100/foot	3 yrs.
	A-2100	Co-60	5	3 yrs.
	A-5772	Co-60	5/foot	3 yrs.
-2	A-2102	Cs-137	1000	3 yrs.
	A-5771	Cs-137	1000/foot	3 yrs.
	A-2100	Co-60	15	3 yrs.
	A-5772	Co-60	15/foot	3 yrs.
-3	A-2104	Cs-137	5000	3 yrs.
	A-5771	Cs-137	1000/foot	3 yrs.
	A-2100	Co-60	60	3 yrs.
	A-5772	Co-60	60/foot	3 yrs.
SHRD	A-2104	Cs-137	20,000	3 yrs.
	A-2100	Co-60	5,000	3 yrs.
SHRH-A	A-2102	Cs-137	750	3 yrs.
	A-2100	Co-60	35	3 yrs.
	A-5771	Cs-137	600/foot	3 yrs.

<u>Device Model Numbers</u>	<u>Source Model Numbers</u>	<u>Isotopes</u>	<u>Maximum Activity Per Source (Millicuries)</u>	<u>Wipe Test Interva</u>
SHRM-PA	A-2102	Cs-137	100	3 yrs.
	A-2100	Co-60	5	3 yrs.
SHRM Series	A-5771	Cs-137	100/foot	3 yrs.
	A-5772	Co-60	10/foot	3 yrs.
SR-1	A-2102	Cs-137	1000	3 yrs.
SR-2	A-2104	Cs-137	5000	3 yrs.
SR-1A	A-2102	Cs-137	1562	3 yrs.
SHWA	A-2102	Cs-137	130	3 yrs.
	A-2100	Co-60	15	3 yrs.
WebART Series	See BAL			
WeighART Series Scale	See: 1) SH-100 or 2) SHWA or 3) D-36698			
D-36698	A-5799	Am-241	500	6 Mos.