

"OFFICIAL RECORD COPY"

MAR 13 1985

West Virginia Institute of Technology
ATTN: Mr. Keith R. Honey, Chairman
Department of Physics
Montgomery, WV 25136

Gentlemen:

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING A MATERIAL LICENSE
APPLICATION, (REFERENCE: 24513; 070-00661)

This refers to your letter dated November 26, 1984, for renewal of License
Number SNM-608. Additional information is needed in support of your request.

By telephone on January 31, 1985, this information was requested; however, our
office has not received a reply to the request as of the date of this letter.
As a reminder, we need the following information:

1. Under Item 11, Calibration of Instruments on your application, the
information was fine for initial calibration of your new equipment; however,
we need to know the frequency these instruments will be calibrated and the
name and address of the service company that will routinely calibrate or
your procedures and standards if you plan to calibrate them.
2. For personnel monitoring, if you use the pocket dosimeters for your primary,
legal personnel exposure record, then you need to submit your provisions for
proper maintenance and calibration of these dosimeters. If you feel that
personnel monitoring is not required in accordance with 10 CFR 20,
20.202(a)(1), please submit an evaluation of the expected exposures from the
proposed uses and demonstrate that no individual is likely to receive a dose
in any calendar quarter in excess of 0.31 rems to the whole body,
blood-forming organs, lens of the eye, or gonads; 4.69 rems to extremities;
or 1.88 rems to the skin from all sources and types of occupational
exposures.

Please provide two copies of your reply and reference Mail Control Number 24513.

You are hereby notified that unless we receive the additional information
requested within thirty (30) days from the date of this notice, we will consider
your application as having been abandoned by you. This action is without
prejudice to the resubmission of an application. If you should have any
questions, do not hesitate to call me at (404) 221-2675.

Sincerely,

Carol A. Connell
Radiation Specialist
Nuclear Materials Safety Section

RII
JPPotter
3/13/85

8505290513 850508
REG2 LIC70
SNM-0608 PDR

245-661

CONVERSATION RECORD

TIME

~4:00 P.M.

DATE

1/31/85

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

ROUTING

NAME/SYMBOL INT

Location of Visit/Conference.

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Office, dept., bureau, etc.)

TELEPHONE NO.

DRS. Honey & Wilson

WEST VA.

(304)

INST. OF TECHNOLOGY

442-3132

SUBJECT

CHARIFICATION ON PERSONNEL DOSIMETRY,

USE FREQUENCY & TIME INT. FOR CALIBRATION OF

SUMMARY

SURVEY INSTRUMENTS

DRS. Honey & Wilson STATED THAT DUE TO INFREQUENT OR FLUCTUATION IN USE OF SOURCES THEY DID NOT THINK THEY NEEDED FILM BADGES OR 3RD PARTY CALIB. OF DOSIMETERS. THEY WILL BE SENDING JUSTIFICATION FOR NOT USING, ALONG WITH TIME INTERVAL THAT SURVEY EQUIP. WILL BE CALIBRATED, IN WRITING WITHIN 30 DAYS.

ACTION REQUIRED

WAIT FOR WRITTEN JUSTIFICATION & COMMITMENT.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

C. CONNELL

C. Connell

1-31-85

ACTION TAKEN

SIGNATURE

TITLE

DATE