

SEP 30 1992

Docket No. 030-32688
Control No. 116236

Technion, Incorporated
ATTN: Susan Baturay
President
250 Delawanna Avenue
Clifton, New Jersey 07014

Dear Ms. Baturay:

This is in reference to your application dated February 20, 1992 for a byproduct material license. In order to continue our review, we need the following additional information:

1. On December 19, 1991 this office sent you a Notice of Violation concerning the fact that you failed file for renewal of NRC License No. 29-20937-01 prior to its expiration. The Notice of Violation directed you to include in your response to it, a statement confirming that you have placed all Shimadzu Scientific Model EDC-7 detector cells which you possess in storage and describing the controls that will be exercised over the storage area, i.e., that the storage area will be locked at all times when not attended. To date, we have not received a response from you which includes this information. Please provide this response so that we can continue review of your license application.
2. Your license application does not describe your program for leak testing your detector cells. Please confirm that detector cells will be leak tested every six months or other interval specified by the manufacturer, and provide the name of the organization which analyzes the leak test samples.
3. Please describe the measures you will take to ensure that gas chromatography devices which contain licensed detector cells and are not in storage will be properly secured against unauthorized removal.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 116236. The reviewer for this licensing action is Keith D. Brown. If you have any technical questions regarding this deficiency letter please call the reviewer at (215) 337-5048.

9212080215 920930
PDR ADDCK 03029031
C PDR

Technion, Incorporated

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,

Francis M. Costello, Chief
Industrial Applications Section
Division of Radiation Safety
and Safeguards