

TECHNION INC.

250 Delawanna Avenue
Clifton, New Jersey 07014
(201) 759-0228
Fax: (201) 773-4788

Francis M. Costello
United States
Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, Pennsylvania 19406-1415

October 21, 1992

Docket No. 030-32688
Control No. 116236

Re: Application dated 02/20/92 (Lic. No. 29-20937-01)

Dear Mr./Ms. Costello:

In accordance with your letter dated September 30, 1992.
Please find below the additional information you requested:

1. The Shimadzu Scientific Model ECD-7 detector cells (instrument) is placed in storage and it is locked when unattended. Only authorized personnel has access to that storage area. This includes the Laboratory Director, the Laboratory Manager, the Laboratory Supervisor and myself (the Acting Safety Officer).
2. The detector cells are leak tested every six months or according to manufacturer's suggestion. The previous organization that analyzed the leak test samples is Medical Industrial Systems Consultants (licensed under MD-03-02201). The current organization that analyzes our leak test samples is National Leak Test Center (NLTC Industries) located in Tonawanda, New York.
3. All gas chromatography devises which contain licensed detector cells which are currently in use are in one room (GC Room). The room is locked and the equipment is used by authorized laboratory personnel only and under the supervision of the Laboratory Manager.

Should you have any questions, feel free to contact me
at the above number.

Thank You.

Sincerely,

Mara Baturay
Mara Baturay,
Assistant Manager
Safety Officer

MB/vs 023

9212080222 921021
PDR ADDCK 03029031
C PDR

116236

OCT 26 1992