

25  
FEB 04 1993

Docket Nos. 50-352  
50-353

Mr. D. M. Smith  
Senior Vice President -  
Nuclear  
Philadelphia Electric Company  
Nuclear Group Headquarters  
Correspondence Control Desk  
P. O. Box 195  
Wayne, PA 19087-0195

Dear Mr. Smith:

Subject: Combined Inspection Nos. 50-352/93-01 and 50-353/93-01

This letter refers to the safety inspection conducted by Mr. J. Kottan and Ms. N. McNamara of this office on January 11-15, 1993, of activities authorized by NRC License Nos. NPF-39 and NPF-85 at the Limerick Generating Station, Units 1 & 2, Limerick, Pennsylvania, and to the discussions of our findings held by Mr. Kottan and Ms. McNamara with Mr. J. Doering and other members of your staff at the conclusion of the inspection.

Areas examined during this inspection are important to public health and safety, and are described in the NRC Region I Inspection Report which is enclosed with this letter. Within these areas, the inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspectors.

In addition, our Mobile Laboratory was brought to your site and used by the inspectors to make certain independent measurements. The basic purpose of these independent measurements is to verify your capability for analyzing radioactive effluents, and to achieve and maintain comparable methods of analyses between your facility and the NRC. The data from these measurements are included in this report with the exception of the wet chemistry results, which will be documented in a subsequent inspection report.

9302160011 930204  
PDR ADOCK 05000352  
G PDR

OFFICIAL RECORD COPY G:LIM9301.CIR February 4, 1993

IE 06

FEB 04 1993

Philadelphia Electric Company

2

Within the scope of this inspection, you had in place effective programs for measuring radioactivity in process and effluent samples and for measuring chemical parameters in plant systems samples. No violations of regulatory requirements were observed and no response to this letter is required.

Sincerely,

Original Signed By:  
James H. Joyner

James H. Joyner, Chief  
Facilities Radiological Safety  
and Safeguards Branch  
Division of Radiation Safety  
and Safeguards

Enclosure: NRC Region I Combined Inspection Report No. 50-352/93-01 and 50-353/93-01

cc w/encl:

R. Charles, Chairman, Nuclear Review Board  
D. R. Helwig, Vice President - Limerick Generating Station  
G. J. Beck, Manager - Licensing Section  
G. Madsen, Regulatory Engineer - Limerick Generating Station  
Secretary, Nuclear Committee of the Board  
Public Document Room (PDR)  
Local Public Document Room (LPDR)  
Nuclear Safety Information Center (NSIC)  
K. Abraham, PAO (2)  
NRC Resident Inspector  
Commonwealth of Pennsylvania

FEB 04 1993

Philadelphia Electric Company

3

bcc w/encl:

Region I Docket Room (with concurrences)

E. Wenzinger, DRP


C. Anderson, DRP

N. Perry, FRI, Yankee Rowe


DRS/EB SALP Coordinator

V. McCree, OEDO


F. Rinaldi, NRR

  
RI:DRSS  
McNamara/tmh


02/4/93

  
RI:DRSS  
Kottan

02/4/93

  
RI:DRSS  
Miller

02/4/93

  
RI:DRSS  
Loyer

02/4/93