

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

37-08784-01

Docket or Reference number

030-06194

Amendment No. 09

Dalare Associates, Inc.
217 South 24th Street
Philadelphia, Pennsylvania 19103

In accordance with letter dated February 22, 1985, License Number 37-08784-01 is amended as follows:

Condition 14. is amended to read:

14. A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed six months; except those detectors as specified by the manufacturer and specifically authorized by the Commission or an Agreement state may be leak tested at intervals not to exceed three years. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the surfaces of the device in which the foil is mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the foil from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the U.S. Nuclear Regulatory Commission, Region I, 631 Park Avenue, King of Prussia, Pennsylvania 19406, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the manufacturer's leak test kit for analysis by Tracor, Inc. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.

B505300593 B50517
REG1 LIC30
37-08784-01 PDR

For the U.S. Nuclear Regulatory Commission

Original Signed By:
Edwin A. Wurtz

By

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

Date MAY 13 1985

"OFFICIAL RECORD COPY"

ML10