

JUN 25 1985

Good Samaritan Medical Center
ATTN: Lutheran Campus
John Schwartz
Chief Operating Officer
620 North 19th Street
Milwaukee, WI 53233

Gentlemen:

Enclosed in Amendment No. 38 to your byproduct material license. As per a telephone conversation between Mr. Schwartz and me on June 14, 1985, we have amended your license in its entirety as a renewal for a five year period (i.e., the appropriate fee was paid and all your procedures were changed as per your March 25, 1985 application) instead of amending it for each request in letters dated November 16, 1984, January 7, 1985, January 15, 1985, and March 25, 1985.

Please note the following:

1. We have not added the physicians (Drs. Whalen, Veluvolu, and Goldstein) as requested. In order to add Drs. Whalen and Veluvolu, resubmit Supplement A with each physician's name in box 1. Both forms sent to us on April 3, 1984 contained Dr. Collier's name. This makes it unclear as to which physician's training is listed on the form.

In order to add Dr. Goldstein to utilize a strontium-90 applicator as requested in your November 16, 1984, application, you will need to resubmit documentation of his active participation using the applicator within the last 5 years. You may wish to have him work under the supervision of an authorized user on your license for a period of time and this could be documented as experience within the past five years. This documentation should be signed and dated by the preceptor. Upon receipt of this documentation, we will add the strontium-90 applicator in accordance with procedures stated in your November 16, 1984 letter.

2. Since you have deleted Drs. Norlund, Walker and Falich, and they were the only authorized users for Group VI material, the authorized use condition (Item 9.E.) has been changed to read "for storage only". At such time as you are able to add authorized physicians for this use, we will reinstate Group VI for active use.

3. We have removed the transport condition from your license since your application states there will be no transfer of material between your locations of use. If this situation changes, you will need to amend your license accordingly.
4. As discussed with Todd Kranpitz via phone on May 30, 1985, we are not adding the letter dated May 22, 1985 to your tie-down condition. Our understanding is that you will run the xenon exhaust fan in the scanning room for a minimum of five minutes per patient exam as stated in your application dated March 25, 1985.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.

7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
Bruce S. Mallett, Ph.D., Chief
Materials Licensing Section

Enclosures:

1. Amendment No. 38
2. 10 CFR Parts 19, 20, 35

RIII

Bsm 6/19/85
Mallett/as
06/18/85