

ENCLOSURE

NOTICE OF VIOLATION

Department of the Army

License No. 16-03657-01

The following violations were identified during an inspection conducted on April 25, 1985. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

1. License Condition 17 requires that licensed material be used in accordance with statements and procedures contained in your application dated December 29, 1983. One procedure in your application requires that radiation level surveys be conducted daily in the isotopes laboratory.

Contrary to the above, radiation level surveys in the isotopes laboratory had been conducted every other day rather than daily.

This is a Severity Level IV violation (Supplement VI).

2. License Condition 17 requires that licensed material be used in accordance with statements and procedures contained in your application dated December 29, 1983. One procedure in your application requires that the linearity of your dose calibrator be tested at quarterly intervals.

Contrary to the above, the linearity of your dose calibrator had not been tested in the first quarter of 1985.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: MAY 17 1985

8507110273 850606
REG2 LIC30
16-03657-01 PDR